

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101023\
 Data File : FB030402.D
 Signal(s) : FID2B.CH
 Acq On : 10 Oct 2023 10:17
 Operator : YP/AJ
 Sample : 20 GRO STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 20 GRO STD

Integration File: Calibration.e
 Quant Time: Oct 10 10:29:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101023.M
 Quant Title :
 QLast Update : Tue Oct 10 10:27:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.784	240211	20.000 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.704	399451	30.000 ng/ml
2) t 2,2,4-Trimethylpentane	7.409	452450	30.000 ng/ml
3) t n-Heptane	7.744	147985	10.000 ng/ml
4) t Benzene	7.880	198825	10.000 ng/ml
6) t Toluene	10.612	577649	30.000 ng/ml
7) t Ethylbenzene	13.049	186131	10.000 ng/ml
8) t m-Xylene	13.182	370393	20.000 ng/ml
9) t O-Xylene	13.908	362852	20.000 ng/ml
10) t 1,2,4-Trimethylbenzene	16.183	371535	20.000 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101023\
Data File : FB030402.D
Signal(s) : FID2B.CH
Acq On : 10 Oct 2023 10:17
Operator : YP/AJ
Sample : 20 GRO STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
20 GRO STD

Integration File: Calibration.e
Quant Time: Oct 10 10:29:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101023.M
Quant Title :
QLast Update : Tue Oct 10 10:27:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.704 min
Delta R.T.: 0.000 min
Response: 399451
Conc: 30.00 ng/ml

Instrument :
FID_B
ClientSampleId :
20 GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.409 min
Delta R.T.: 0.000 min
Response: 452450
Conc: 30.00 ng/ml




#3 n-Heptane

R.T.: 7.744 min
Delta R.T.: 0.000 min
Response: 147985
Conc: 10.00 ng/ml



#4 Benzene


R.T.: 7.880 min
Delta R.T.: 0.000 min
Response: 198825
Conc: 10.00 ng/ml



#5 AAA-TFT


R.T.: 8.784 min
Delta R.T.: 0.000 min
Response: 240211
Conc: 20.00 ng/ml

Instrument :
FID_B
ClientSampleId :
20 GRO STD




#6 Toluene

R.T.: 10.612 min
Delta R.T.: 0.000 min
Response: 577649
Conc: 30.00 ng/ml




#7 Ethylbenzene

R.T.: 13.049 min
Delta R.T.: 0.000 min
Response: 186131
Conc: 10.00 ng/ml



#8 m-Xylene


R.T.: 13.182 min
Delta R.T.: 0.000 min
Response: 370393
Conc: 20.00 ng/ml



#9 O-Xylene

R.T.: 13.908 min
Delta R.T.: 0.000 min
Response: 362852
Conc: 20.00 ng/ml

Instrument :
FID_B
ClientSampleId :
20 GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.183 min
Delta R.T.: 0.000 min
Response: 371535
Conc: 20.00 ng/ml