

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB101024\  
 Data File : FB030992.D  
 Signal(s) : FID2B.CH  
 Acq On : 10 Oct 2024 14:31  
 Operator : YP/AJ  
 Sample : BSF1010W1  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 FID\_B  
 ClientSampleId :

Integration File: Calibration.e  
 Quant Time: Oct 11 04:32:44 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB100424.M  
 Quant Title :  
 QLast Update : Fri Oct 04 11:53:51 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 5 g/ml  
 Signal Phase : RTX-502.2  
 Signal Info : 60mx0.53mmx3.00um

| Compound                     | R.T.   | Response | Conc Units   |
|------------------------------|--------|----------|--------------|
| -----                        |        |          |              |
| System Monitoring Compounds  |        |          |              |
| 5) s AAA-TFT                 | 8.795  | 596328   | 19.893 ng/ml |
| Target Compounds             |        |          |              |
| 1) t 2-Methylpentane         | 4.727  | 983838   | 27.265 ng/ml |
| 2) t 2,2,4-Trimethylpentane  | 7.428  | 1287032  | 27.008 ng/ml |
| 3) t n-Heptane               | 7.757  | 396538   | 8.792 ng/ml  |
| 4) t Benzene                 | 7.897  | 505164   | 9.191 ng/ml  |
| 6) t Toluene                 | 10.624 | 1452625  | 26.891 ng/ml |
| 7) t Ethylbenzene            | 13.060 | 431820   | 9.110 ng/ml  |
| 8) t m-Xylene                | 13.195 | 938432   | 17.989 ng/ml |
| 9) t O-Xylene                | 13.923 | 874677   | 17.981 ng/ml |
| 10) t 1,2,4-Trimethylbenzene | 16.198 | 653983   | 19.958 ng/ml |
| -----                        |        |          |              |

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB101024\  
Data File : FB030992.D  
Signal(s) : FID2B.CH  
Acq On : 10 Oct 2024 14:31  
Operator : YP/AJ  
Sample : BSF1010W1  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
FID\_B  
ClientSampleId :

Integration File: Calibration.e  
Quant Time: Oct 11 04:32:44 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB100424.M  
Quant Title :  
QLast Update : Fri Oct 04 11:53:51 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 5 g/ml  
Signal Phase : RTX-502.2  
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.727 min  
Delta R.T.: -0.001 min  
Response: 983838  
Conc: 27.26 ng/ml

Instrument :  
FID\_B  
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.428 min  
Delta R.T.: 0.000 min  
Response: 1287032  
Conc: 27.01 ng/ml




#3 n-Heptane

R.T.: 7.757 min  
Delta R.T.: 0.000 min  
Response: 396538  
Conc: 8.79 ng/ml



#4 Benzene


R.T.: 7.897 min  
Delta R.T.: 0.000 min  
Response: 505164  
Conc: 9.19 ng/ml



#5 AAA-TFT


R.T.: 8.795 min  
Delta R.T.: 0.000 min  
Response: 596328  
Conc: 19.89 ng/ml

Instrument :  
FID\_B  
ClientSampleId :




#6 Toluene

R.T.: 10.624 min  
Delta R.T.: 0.000 min  
Response: 1452625  
Conc: 26.89 ng/ml




#7 Ethylbenzene

R.T.: 13.060 min  
Delta R.T.: 0.000 min  
Response: 431820  
Conc: 9.11 ng/ml



#8 m-Xylene


R.T.: 13.195 min  
Delta R.T.: 0.000 min  
Response: 938432  
Conc: 17.99 ng/ml



#9 O-Xylene

R.T.: 13.923 min  
Delta R.T.: 0.000 min  
Response: 874677  
Conc: 17.98 ng/ml

Instrument :  
FID\_B  
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.198 min  
Delta R.T.: 0.000 min  
Response: 653983  
Conc: 19.96 ng/ml