

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101118\
Data File : FB018206.D
Signal(s) : FID2B.CH
Acq On : 11 Oct 2018 20:09
Operator : MJ/SJ
Sample : 5 GRO STD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 12 07:59:10 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
Quant Title :
QLast Update : Fri Oct 12 07:57:44 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.119	43754	5.455 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.965	114886	7.896 ng/ml
2) t 2,2,4-Trimethylpentane	7.728	108302	7.524 ng/ml
3) t n-Heptane	8.067	34405	2.465 ng/ml
4) t Benzene	8.207	46732	2.608 ng/ml
6) t Toluene	10.963	119148	8.000 ng/ml
7) t Ethylbenzene	13.411	44227	2.809 ng/ml
8) t m-Xylene	13.546	92403	5.555 ng/ml
9) t O-Xylene	14.278	99954	5.681 ng/ml
10) t 1,2,4-Trimethylbenzene	16.560	114034	5.652 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101118\
Data File : FB018206.D
Signal(s) : FID2B.CH
Acq On : 11 Oct 2018 20:09
Operator : MJ/SJ
Sample : 5 GRO STD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 12 07:59:10 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
Quant Title :
QLast Update : Fri Oct 12 07:57:44 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.965 min
Delta R.T.: 0.000 min
Response: 114886
Conc: 7.90 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.728 min
Delta R.T.: 0.000 min
Response: 108302
Conc: 7.52 ng/ml




#3 n-Heptane

R.T.: 8.067 min
Delta R.T.: 0.000 min
Response: 34405
Conc: 2.47 ng/ml




#4 Benzene

R.T.: 8.207 min
Delta R.T.: 0.000 min
Response: 46732
Conc: 2.61 ng/ml




#5 AAA-TFT

R.T.: 9.119 min
Delta R.T.: 0.006 min
Response: 43754
Conc: 5.46 ng/ml




#6 Toluene

R.T.: 10.963 min
Delta R.T.: 0.001 min
Response: 119148
Conc: 8.00 ng/ml




#7 Ethylbenzene

R.T.: 13.411 min
Delta R.T.: 0.000 min
Response: 44227
Conc: 2.81 ng/ml




#8 m-Xylene

R.T.: 13.546 min
Delta R.T.: 0.002 min
Response: 92403
Conc: 5.56 ng/ml



#9 O-Xylene

R.T.: 14.278 min
Delta R.T.: 0.002 min
Response: 99954
Conc: 5.68 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.560 min
Delta R.T.: 0.000 min
Response: 114034
Conc: 5.65 ng/ml