

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101118\
 Data File : FB018210.D
 Signal(s) : FID2B.CH
 Acq On : 11 Oct 2018 22:10
 Operator : MJ/SJ
 Sample : 100 GRO STD
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Oct 12 07:59:22 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
 Quant Title :
 QLast Update : Fri Oct 12 07:57:44 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.111	851069	106.115 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.960	2460210	169.089 ng/ml
2) t 2,2,4-Trimethylpentane	7.727	2411695	167.541 ng/ml
3) t n-Heptane	8.067	766416	54.911 ng/ml
4) t Benzene	8.206	1001647	55.908 ng/ml
6) t Toluene	10.962	2380493	159.834 ng/ml
7) t Ethylbenzene	13.411	798252	50.698 ng/ml
8) t m-Xylene	13.544	1688168	101.496 ng/ml
9) t O-Xylene	14.276	1817768	103.316 ng/ml
10) t 1,2,4-Trimethylbenzene	16.559	2035517	100.896 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101118\
Data File : FB018210.D
Signal(s) : FID2B.CH
Acq On : 11 Oct 2018 22:10
Operator : MJ/SJ
Sample : 100 GRO STD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 12 07:59:22 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
Quant Title :
QLast Update : Fri Oct 12 07:57:44 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.960 min
Delta R.T.: -0.005 min
Response: 2460210
Conc: 169.09 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.727 min
Delta R.T.: 0.000 min
Response: 2411695
Conc: 167.54 ng/ml




#3 n-Heptane

R.T.: 8.067 min
Delta R.T.: 0.000 min
Response: 766416
Conc: 54.91 ng/ml




#4 Benzene

R.T.: 8.206 min
Delta R.T.: 0.000 min
Response: 1001647
Conc: 55.91 ng/ml




#5 AAA-TFT

R.T.: 9.111 min
Delta R.T.: -0.001 min
Response: 851069
Conc: 106.12 ng/ml




#6 Toluene

R.T.: 10.962 min
Delta R.T.: 0.000 min
Response: 2380493
Conc: 159.83 ng/ml




#7 Ethylbenzene

R.T.: 13.411 min
Delta R.T.: 0.000 min
Response: 798252
Conc: 50.70 ng/ml




#8 m-Xylene

R.T.: 13.544 min
Delta R.T.: 0.000 min
Response: 1688168
Conc: 101.50 ng/ml



#9 O-Xylene

R.T.: 14.276 min
Delta R.T.: 0.000 min
Response: 1817768
Conc: 103.32 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.559 min
Delta R.T.: 0.000 min
Response: 2035517
Conc: 100.90 ng/ml