

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101421\
 Data File : FB026286.D
 Signal(s) : FID2B.CH
 Acq On : 14 Oct 2021 14:59
 Operator : AJ/MA
 Sample : M4164-01MSD
 Misc : 5.00G\5.00 ML DI WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BU-03-100821MSD

Integration File: SAMPLE.e
 Quant Time: Oct 14 15:21:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100821.M
 Quant Title :
 QLast Update : Fri Oct 08 14:30:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.848	228801	15.183 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.846	329523	18.930 ng/ml
2) t 2,2,4-Trimethylpentane	7.521	330998	16.024 ng/ml
3) t n-Heptane	7.821	110204	4.704 ng/ml
4) t Benzene	7.960	205578	7.857 ng/ml
6) t Toluene	10.662	567345	21.899 ng/ml
7) t Ethylbenzene	13.086	163068	7.140 ng/ml
8) t m-Xylene	13.218	321466	14.331 ng/ml
9) t O-Xylene	13.944	308808	14.597 ng/ml
10) t 1,2,4-Trimethylbenzene	16.216	288448	12.396 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101421\
Data File : FB026286.D
Signal(s) : FID2B.CH
Acq On : 14 Oct 2021 14:59
Operator : AJ/MA
Sample : M4164-01MSD
Misc : 5.00G\5.00 ML DI WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BU-03-100821MSD

Integration File: SAMPLE.e
Quant Time: Oct 14 15:21:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100821.M
Quant Title :
QLast Update : Fri Oct 08 14:30:36 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

