

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101421\
 Data File : FB026288.D
 Signal(s) : FID2B.CH
 Acq On : 14 Oct 2021 16:21
 Operator : AJ/MA
 Sample : 20 PPB GRO STD
 Misc : 5.01G\5.00 ML DI WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 20 PPB GRO STD

Integration File: SAMPLE.e
 Quant Time: Oct 14 23:33:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100821.M
 Quant Title :
 QLast Update : Fri Oct 08 14:30:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.854	283809	18.833 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.869	489915	28.144 ng/ml
2) t 2,2,4-Trimethylpentane	7.542	605562	29.316 ng/ml
3) t n-Heptane	7.830	191342	8.167 ng/ml
4) t Benzene	7.969	229400	8.767 ng/ml
6) t Toluene	10.665	672691	25.965 ng/ml
7) t Ethylbenzene	13.087	206075	9.023 ng/ml
8) t m-Xylene	13.219	413975	18.455 ng/ml
9) t O-Xylene	13.944	395131	18.678 ng/ml
10) t 1,2,4-Trimethylbenzene	16.215	422445	18.154 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101421\
Data File : FB026288.D
Signal(s) : FID2B.CH
Acq On : 14 Oct 2021 16:21
Operator : AJ/MA
Sample : 20 PPB GRO STD
Misc : 5.01G\5.00 ML DI WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD

Integration File: SAMPLE.e
Quant Time: Oct 14 23:33:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100821.M
Quant Title :
QLast Update : Fri Oct 08 14:30:36 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

