

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB101920\  
 Data File : FB025106.D  
 Signal(s) : FID2B.CH  
 Acq On : 19 Oct 2020 22:43  
 Operator : DD/AJ  
 Sample : 20 PPB GRO STD  
 Misc :  
 ALS Vial : 24 Sample Multiplier: 1

Integration File: Calibration.e  
 Quant Time: Oct 20 01:02:30 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB101920.M  
 Quant Title :  
 QLast Update : Mon Oct 19 13:01:11 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 5 g/ml  
 Signal Phase : RTX-502.2  
 Signal Info : 60mx0.53mmx3.00um

| Compound                     | R.T.   | Response | Conc Units   |
|------------------------------|--------|----------|--------------|
| -----                        |        |          |              |
| System Monitoring Compounds  |        |          |              |
| 5) s AAA-TFT                 | 8.798  | 354050   | 16.585 ng/ml |
| Target Compounds             |        |          |              |
| 1) t 2-Methylpentane         | 4.688  | 746221   | 24.535 ng/ml |
| 2) t 2,2,4-Trimethylpentane  | 7.416  | 786072   | 25.233 ng/ml |
| 3) t n-Heptane               | 7.757  | 255885   | 8.323 ng/ml  |
| 4) t Benzene                 | 7.892  | 348696   | 9.278 ng/ml  |
| 6) t Toluene                 | 10.630 | 983164   | 27.587 ng/ml |
| 7) t Ethylbenzene            | 13.069 | 299345   | 9.196 ng/ml  |
| 8) t m-Xylene                | 13.202 | 593140   | 18.495 ng/ml |
| 9) t O-Xylene                | 13.929 | 585075   | 18.979 ng/ml |
| 10) t 1,2,4-Trimethylbenzene | 16.207 | 488824   | 18.723 ng/ml |
| -----                        |        |          |              |


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB101920\  
Data File : FB025106.D  
Signal(s) : FID2B.CH  
Acq On : 19 Oct 2020 22:43  
Operator : DD/AJ  
Sample : 20 PPB GRO STD  
Misc :  
ALS Vial : 24 Sample Multiplier: 1

Integration File: Calibration.e  
Quant Time: Oct 20 01:02:30 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB101920.M  
Quant Title :  
QLast Update : Mon Oct 19 13:01:11 2020  
Response via : Initial Calibration  
Integrator: ChemStation


Volume Inj. : 5 g/ml  
Signal Phase : RTX-502.2  
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.688 min  
Delta R.T.: 0.000 min  
Response: 746221  
Conc: 24.53 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min  
Delta R.T.: -0.002 min  
Response: 786072  
Conc: 25.23 ng/ml




#3 n-Heptane

R.T.: 7.757 min  
Delta R.T.: 0.000 min  
Response: 255885  
Conc: 8.32 ng/ml




#4 Benzene

R.T.: 7.892 min  
Delta R.T.: -0.002 min  
Response: 348696  
Conc: 9.28 ng/ml




#5 AAA-TFT

R.T.: 8.798 min  
Delta R.T.: -0.002 min  
Response: 354050  
Conc: 16.58 ng/ml




#6 Toluene

R.T.: 10.630 min  
Delta R.T.: -0.001 min  
Response: 983164  
Conc: 27.59 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min  
Delta R.T.: -0.001 min  
Response: 299345  
Conc: 9.20 ng/ml




#8 m-Xylene

R.T.: 13.202 min  
Delta R.T.: -0.001 min  
Response: 593140  
Conc: 18.49 ng/ml



#9 O-Xylene

R.T.: 13.929 min  
Delta R.T.: -0.001 min  
Response: 585075  
Conc: 18.98 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.207 min  
Delta R.T.: -0.001 min  
Response: 488824  
Conc: 18.72 ng/ml