

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB102220\
 Data File : FB025126.D
 Signal(s) : FID2B.CH
 Acq On : 22 Oct 2020 9:40
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 20 PPB GRO STD

Integration File: Calibration.e
 Quant Time: Oct 23 00:39:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101920.M
 Quant Title :
 QLast Update : Mon Oct 19 13:01:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.795	391182	18.324 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.687	849506	27.931 ng/ml
2) t 2,2,4-Trimethylpentane	7.413	865438	27.781 ng/ml
3) t n-Heptane	7.754	285798	9.296 ng/ml
4) t Benzene	7.889	358841	9.548 ng/ml
6) t Toluene	10.627	1002592	28.132 ng/ml
7) t Ethylbenzene	13.067	301873	9.274 ng/ml
8) t m-Xylene	13.200	607047	18.928 ng/ml
9) t O-Xylene	13.927	591884	19.200 ng/ml
10) t 1,2,4-Trimethylbenzene	16.205	481841	18.455 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB102220\
Data File : FB025126.D
Signal(s) : FID2B.CH
Acq On : 22 Oct 2020 9:40
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD

Integration File: Calibration.e
Quant Time: Oct 23 00:39:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101920.M
Quant Title :
QLast Update : Mon Oct 19 13:01:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.687 min
Delta R.T.: 0.000 min
Response: 849506
Conc: 27.93 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.413 min
Delta R.T.: -0.005 min
Response: 865438
Conc: 27.78 ng/ml




#3 n-Heptane

R.T.: 7.754 min
Delta R.T.: -0.004 min
Response: 285798
Conc: 9.30 ng/ml



#4 Benzene


R.T.: 7.889 min
Delta R.T.: -0.004 min
Response: 358841
Conc: 9.55 ng/ml



#5 AAA-TFT


R.T.: 8.795 min
Delta R.T.: -0.005 min
Response: 391182
Conc: 18.32 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#6 Toluene

R.T.: 10.627 min
Delta R.T.: -0.004 min
Response: 1002592
Conc: 28.13 ng/ml




#7 Ethylbenzene

R.T.: 13.067 min
Delta R.T.: -0.004 min
Response: 301873
Conc: 9.27 ng/ml



#8 m-Xylene


R.T.: 13.200 min
Delta R.T.: -0.004 min
Response: 607047
Conc: 18.93 ng/ml



#9 O-Xylene

R.T.: 13.927 min
Delta R.T.: -0.004 min
Response: 591884
Conc: 19.20 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.205 min
Delta R.T.: -0.004 min
Response: 481841
Conc: 18.46 ng/ml