

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB102318\
Data File : FB018314.D
Signal(s) : FID2B.CH
Acq On : 23 Oct 2018 8:54
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 24 08:36:07 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
Quant Title :
QLast Update : Fri Oct 12 08:04:31 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.107	170662	19.815 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.959	478555	31.290 ng/ml
2) t 2,2,4-Trimethylpentane	7.722	478258	32.366 ng/ml
3) t n-Heptane	8.062	158272	11.169 ng/ml
4) t Benzene	8.201	201393	10.811 ng/ml
6) t Toluene	10.955	491162	32.227 ng/ml
7) t Ethylbenzene	13.405	168325	10.609 ng/ml
8) t m-Xylene	13.538	358730	21.560 ng/ml
9) t O-Xylene	14.270	376049	21.091 ng/ml
10) t 1,2,4-Trimethylbenzene	16.553	444052	21.828 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB102318\
Data File : FB018314.D
Signal(s) : FID2B.CH
Acq On : 23 Oct 2018 8:54
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 24 08:36:07 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
Quant Title :
QLast Update : Fri Oct 12 08:04:31 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.959 min
Delta R.T.: -0.006 min
Response: 478555
Conc: 31.29 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.722 min
Delta R.T.: -0.007 min
Response: 478258
Conc: 32.37 ng/ml




#3 n-Heptane

R.T.: 8.062 min
Delta R.T.: -0.005 min
Response: 158272
Conc: 11.17 ng/ml




#4 Benzene

R.T.: 8.201 min
Delta R.T.: -0.006 min
Response: 201393
Conc: 10.81 ng/ml




#5 AAA-TFT

R.T.: 9.107 min
Delta R.T.: -0.006 min
Response: 170662
Conc: 19.81 ng/ml




#6 Toluene

R.T.: 10.955 min
Delta R.T.: -0.007 min
Response: 491162
Conc: 32.23 ng/ml




#7 Ethylbenzene

R.T.: 13.405 min
Delta R.T.: -0.007 min
Response: 168325
Conc: 10.61 ng/ml




#8 m-Xylene

R.T.: 13.538 min
Delta R.T.: -0.006 min
Response: 358730
Conc: 21.56 ng/ml



#9 O-Xylene

R.T.: 14.270 min
Delta R.T.: -0.007 min
Response: 376049
Conc: 21.09 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.553 min
Delta R.T.: -0.006 min
Response: 444052
Conc: 21.83 ng/ml