

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB102319\
 Data File : FB022815.D
 Signal(s) : FID2B.CH
 Acq On : 23 Oct 2019 19:09
 Operator : DD/AJ
 Sample : BSF1023W2
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Oct 24 00:40:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101619.M
 Quant Title :
 QLast Update : Wed Oct 16 15:04:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.798	438083	17.313 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.688	836066	24.601 ng/ml
2) t	2,2,4-Trimethylpentane	7.416	918525	25.243 ng/ml
3) t	n-Heptane	7.756	298068	8.169 ng/ml
4) t	Benzene	7.891	392251	8.884 ng/ml
6) t	Toluene	10.629	1140577	26.463 ng/ml
7) t	Ethylbenzene	13.069	362355	8.922 ng/ml
8) t	m-Xylene	13.202	688009	17.882 ng/ml
9) t	O-Xylene	13.930	666415	18.297 ng/ml
10) t	1,2,4-Trimethylbenzene	16.208	521641	18.306 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB102319\
Data File : FB022815.D
Signal(s) : FID2B.CH
Acq On : 23 Oct 2019 19:09
Operator : DD/AJ
Sample : BSF1023W2
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 24 00:40:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101619.M
Quant Title :
QLast Update : Wed Oct 16 15:04:00 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.688 min
Delta R.T.: -0.002 min
Response: 836066
Conc: 24.60 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min
Delta R.T.: -0.001 min
Response: 918525
Conc: 25.24 ng/ml




#3 n-Heptane

R.T.: 7.756 min
Delta R.T.: -0.001 min
Response: 298068
Conc: 8.17 ng/ml




#4 Benzene

R.T.: 7.891 min
Delta R.T.: 0.000 min
Response: 392251
Conc: 8.88 ng/ml




#5 AAA-TFT

R.T.: 8.798 min
Delta R.T.: 0.000 min
Response: 438083
Conc: 17.31 ng/ml




#6 Toluene

R.T.: 10.629 min
Delta R.T.: 0.000 min
Response: 1140577
Conc: 26.46 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min
Delta R.T.: 0.000 min
Response: 362355
Conc: 8.92 ng/ml




#8 m-Xylene

R.T.: 13.202 min
Delta R.T.: 0.000 min
Response: 688009
Conc: 17.88 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: 0.000 min
Response: 666415
Conc: 18.30 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min
Delta R.T.: -0.002 min
Response: 521641
Conc: 18.31 ng/ml