

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB102920\
 Data File : FB025157.D
 Signal(s) : FID2B.CH
 Acq On : 29 Oct 2020 8:37
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Oct 30 00:47:21 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101920.M
 Quant Title :
 QLast Update : Mon Oct 19 13:01:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.799	452669	21.204 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.688	911032	29.954 ng/ml
2) t	2,2,4-Trimethylpentane	7.417	941457	30.221 ng/ml
3) t	n-Heptane	7.757	310048	10.084 ng/ml
4) t	Benzene	7.892	388086	10.326 ng/ml
6) t	Toluene	10.630	1095074	30.727 ng/ml
7) t	Ethylbenzene	13.071	331832	10.194 ng/ml
8) t	m-Xylene	13.204	663147	20.678 ng/ml
9) t	O-Xylene	13.931	642515	20.843 ng/ml
10) t	1,2,4-Trimethylbenzene	16.209	528719	20.251 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB102920\
Data File : FB025157.D
Signal(s) : FID2B.CH
Acq On : 29 Oct 2020 8:37
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 30 00:47:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101920.M
Quant Title :
QLast Update : Mon Oct 19 13:01:11 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 911032
Conc: 29.95 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.417 min
Delta R.T.: 0.000 min
Response: 941457
Conc: 30.22 ng/ml




#3 n-Heptane

R.T.: 7.757 min
Delta R.T.: 0.000 min
Response: 310048
Conc: 10.08 ng/ml




#4 Benzene

R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 388086
Conc: 10.33 ng/ml




#5 AAA-TFT

R.T.: 8.799 min
Delta R.T.: 0.000 min
Response: 452669
Conc: 21.20 ng/ml




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.000 min
Response: 1095074
Conc: 30.73 ng/ml




#7 Ethylbenzene

R.T.: 13.071 min
Delta R.T.: 0.000 min
Response: 331832
Conc: 10.19 ng/ml




#8 m-Xylene

R.T.: 13.204 min
Delta R.T.: 0.000 min
Response: 663147
Conc: 20.68 ng/ml



#9 O-Xylene

R.T.: 13.931 min
Delta R.T.: 0.000 min
Response: 642515
Conc: 20.84 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.209 min
Delta R.T.: 0.000 min
Response: 528719
Conc: 20.25 ng/ml