

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB103019\
 Data File : FB022854.D
 Signal(s) : FID2B.CH
 Acq On : 30 Oct 2019 15:00
 Operator : DD/AJ
 Sample : K5148-14MS
 Misc : 5.02g/5.00ml DI WATER
 ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Oct 31 00:55:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101619.M
 Quant Title :
 QLast Update : Wed Oct 16 15:04:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	521401	20.605 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	841423	24.759 ng/ml
2) t 2,2,4-Trimethylpentane	7.416	825862	22.696 ng/ml
3) t n-Heptane	7.756	271500	7.441 ng/ml
4) t Benzene	7.890	422674	9.573 ng/ml
6) t Toluene	10.629	6237760	144.726 ng/ml
7) t Ethylbenzene	13.068	863847	21.269 ng/ml
8) t m-Xylene	13.200	2397898	62.322 ng/ml
9) t O-Xylene	13.930	1068235	29.329 ng/ml
10) t 1,2,4-Trimethylbenzene	16.209	486497	17.073 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB103019\
Data File : FB022854.D
Signal(s) : FID2B.CH
Acq On : 30 Oct 2019 15:00
Operator : DD/AJ
Sample : K5148-14MS
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 31 00:55:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101619.M
Quant Title :
QLast Update : Wed Oct 16 15:04:00 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.689 min
Delta R.T.: -0.001 min
Response: 841423
Conc: 24.76 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min
Delta R.T.: -0.001 min
Response: 825862
Conc: 22.70 ng/ml




#3 n-Heptane

R.T.: 7.756 min
Delta R.T.: 0.000 min
Response: 271500
Conc: 7.44 ng/ml




#4 Benzene

R.T.: 7.890 min
Delta R.T.: -0.001 min
Response: 422674
Conc: 9.57 ng/ml




#5 AAA-TFT

R.T.: 8.797 min
Delta R.T.: -0.001 min
Response: 521401
Conc: 20.61 ng/ml




#6 Toluene

R.T.: 10.629 min
Delta R.T.: 0.000 min
Response: 6237760
Conc: 144.73 ng/ml




#7 Ethylbenzene

R.T.: 13.068 min
Delta R.T.: -0.001 min
Response: 863847
Conc: 21.27 ng/ml




#8 m-Xylene

R.T.: 13.200 min
Delta R.T.: -0.003 min
Response: 2397898
Conc: 62.32 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: 0.000 min
Response: 1068235
Conc: 29.33 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.209 min
Delta R.T.: -0.001 min
Response: 486497
Conc: 17.07 ng/ml