

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB103019\
 Data File : FB022861.D
 Signal(s) : FID2B.CH
 Acq On : 30 Oct 2019 18:50
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Oct 31 00:56:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101619.M
 Quant Title :
 QLast Update : Wed Oct 16 15:04:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.798	548673	21.683 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	1034728	30.447 ng/ml
2) t 2,2,4-Trimethylpentane	7.417	1165080	32.019 ng/ml
3) t n-Heptane	7.756	379492	10.401 ng/ml
4) t Benzene	7.890	455533	10.317 ng/ml
6) t Toluene	10.629	1334853	30.971 ng/ml
7) t Ethylbenzene	13.070	424015	10.440 ng/ml
8) t m-Xylene	13.203	800053	20.794 ng/ml
9) t O-Xylene	13.930	755266	20.737 ng/ml
10) t 1,2,4-Trimethylbenzene	16.209	597375	20.964 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB103019\
Data File : FB022861.D
Signal(s) : FID2B.CH
Acq On : 30 Oct 2019 18:50
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 31 00:56:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101619.M
Quant Title :
QLast Update : Wed Oct 16 15:04:00 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.688 min
Delta R.T.: -0.002 min
Response: 1034728
Conc: 30.45 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.417 min
Delta R.T.: 0.000 min
Response: 1165080
Conc: 32.02 ng/ml




#3 n-Heptane

R.T.: 7.756 min
Delta R.T.: 0.000 min
Response: 379492
Conc: 10.40 ng/ml




#4 Benzene

R.T.: 7.890 min
Delta R.T.: -0.001 min
Response: 455533
Conc: 10.32 ng/ml




#5 AAA-TFT

R.T.: 8.798 min
Delta R.T.: 0.000 min
Response: 548673
Conc: 21.68 ng/ml




#6 Toluene

R.T.: 10.629 min
Delta R.T.: 0.000 min
Response: 1334853
Conc: 30.97 ng/ml




#7 Ethylbenzene

R.T.: 13.070 min
Delta R.T.: 0.000 min
Response: 424015
Conc: 10.44 ng/ml




#8 m-Xylene

R.T.: 13.203 min
Delta R.T.: 0.000 min
Response: 800053
Conc: 20.79 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: 0.000 min
Response: 755266
Conc: 20.74 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.209 min
Delta R.T.: 0.000 min
Response: 597375
Conc: 20.96 ng/ml