

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB110118\
 Data File : FB018428.D
 Signal(s) : FID2B.CH
 Acq On : 1 Nov 2018 22:54
 Operator : MJ/SJ
 Sample : J5759-11
 Misc : 5.04g/5.00mL DI Water
 ALS Vial : 11 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Nov 02 04:39:36 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
 Quant Title :
 QLast Update : Fri Oct 12 08:04:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.112	171427	19.904 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.948	24294	1.588 ng/ml
2) t	2,2,4-Trimethylpentane	7.774	2578	0.174 ng/ml
3) t	n-Heptane	8.068	80273	5.664 ng/ml
4) t	Benzene	8.215	10843	0.582 ng/ml
6) t	Toluene	10.977	5472	0.359 ng/ml
7) t	Ethylbenzene	13.412	5679	0.358 ng/ml
9) t	O-Xylene	14.216	977	0.055 ng/ml
10) t	1,2,4-Trimethylbenzene	16.569	5117	0.252 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB110118\
Data File : FB018428.D
Signal(s) : FID2B.CH
Acq On : 1 Nov 2018 22:54
Operator : MJ/SJ
Sample : J5759-11
Misc : 5.04g/5.00mL DI Water
ALS Vial : 11 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 02 04:39:36 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
Quant Title :
QLast Update : Fri Oct 12 08:04:31 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.948 min
Delta R.T.: -0.017 min
Response: 24294
Conc: 1.59 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.774 min
Delta R.T.: 0.046 min
Response: 2578
Conc: 0.17 ng/ml




#3 n-Heptane

R.T.: 8.068 min
Delta R.T.: 0.000 min
Response: 80273
Conc: 5.66 ng/ml




#4 Benzene

R.T.: 8.215 min
Delta R.T.: 0.008 min
Response: 10843
Conc: 0.58 ng/ml




#5 AAA-TFT

R.T.: 9.112 min
Delta R.T.: 0.000 min
Response: 171427
Conc: 19.90 ng/ml




#6 Toluene

R.T.: 10.977 min
Delta R.T.: 0.015 min
Response: 5472
Conc: 0.36 ng/ml




#7 Ethylbenzene

R.T.: 13.412 min
Delta R.T.: 0.000 min
Response: 5679
Conc: 0.36 ng/ml



#9 O-Xylene

R.T.: 14.216 min
Delta R.T.: -0.060 min
Response: 977
Conc: 0.05 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.569 min
Delta R.T.: 0.009 min
Response: 5117
Conc: 0.25 ng/ml