

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\FID_B\Data\FB110118\
 Data File : FB018403.D
 Signal(s) : FID2B.CH
 Acq On : 1 Nov 2018 10:07
 Operator : MJ/SJ
 Sample : BSF1101S1
 Misc : 5.02g/5.00mL DI Water
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Nov 02 04:36:19 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
 Quant Title :
 QLast Update : Fri Oct 12 08:04:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.113	175805	20.412 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.964	442985	28.964 ng/ml
2) t 2,2,4-Trimethylpentane	7.728	459370	31.088 ng/ml
3) t n-Heptane	8.069	149115	10.522 ng/ml
4) t Benzene	8.209	191650	10.288 ng/ml
6) t Toluene	10.963	474264	31.118 ng/ml
7) t Ethylbenzene	13.413	148628	9.368 ng/ml
8) t m-Xylene	13.546	314161	18.881 ng/ml
9) t O-Xylene	14.277	320503	17.976 ng/ml
10) t 1,2,4-Trimethylbenzene	16.560	405565	19.936 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\FID_B\Data\FB110118\
Data File : FB018403.D
Signal(s) : FID2B.CH
Acq On : 1 Nov 2018 10:07
Operator : MJ/SJ
Sample : BSF1101S1
Misc : 5.02g/5.00mL DI Water
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 02 04:36:19 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
Quant Title :
QLast Update : Fri Oct 12 08:04:31 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

