

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB110218\
Data File : FB018490.D
Signal(s) : FID2B.CH
Acq On : 3 Nov 2018 8:41
Operator : MJ/SJ
Sample : J5771-05
Misc : 5.04g/5.00mL DI Water
ALS Vial : 44 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 05 04:39:39 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
Quant Title :
QLast Update : Fri Oct 12 08:04:31 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.111	191966	22.288 ng/ml
Target Compounds				
6) t	Toluene	10.972	4046	0.265 ng/ml
8) t	m-Xylene	13.541	2666	0.160 ng/ml
9) t	O-Xylene	14.240	434	0.024 ng/ml
10) t	1,2,4-Trimethylbenzene	16.561	7427	0.365 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB110218\
Data File : FB018490.D
Signal(s) : FID2B.CH
Acq On : 3 Nov 2018 8:41
Operator : MJ/SJ
Sample : J5771-05
Misc : 5.04g/5.00mL DI Water
ALS Vial : 44 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 05 04:39:39 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB101218.M
Quant Title :
QLast Update : Fri Oct 12 08:04:31 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT

R.T.: 9.111 min
Delta R.T.: -0.001 min
Response: 191966
Conc: 22.29 ng/ml




#6 Toluene

R.T.: 10.972 min
Delta R.T.: 0.010 min
Response: 4046
Conc: 0.27 ng/ml




#8 m-Xylene

R.T.: 13.541 min
Delta R.T.: -0.004 min
Response: 2666
Conc: 0.16 ng/ml



#9 O-Xylene

R.T.: 14.240 min
Delta R.T.: -0.037 min
Response: 434
Conc: 0.02 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.561 min
Delta R.T.: 0.002 min
Response: 7427
Conc: 0.37 ng/ml