

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB110520\
 Data File : FB025196.D
 Signal(s) : FID2B.CH
 Acq On : 5 Nov 2020 13:47
 Operator : DD/AJ
 Sample : L4621-05MS
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 PAV-B8-110320-10-20MS

Integration File: Calibration.e
 Quant Time: Nov 05 15:20:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110320.M
 Quant Title :
 QLast Update : Wed Nov 04 08:05:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	445869	19.747 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	892958	27.665 ng/ml
2) t 2,2,4-Trimethylpentane	7.414	881266	26.562 ng/ml
3) t n-Heptane	7.755	295914	9.009 ng/ml
4) t Benzene	7.891	389879	9.806 ng/ml
6) t Toluene	10.629	1096337	29.321 ng/ml
7) t Ethylbenzene	13.068	352991	10.279 ng/ml
8) t m-Xylene	13.201	684559	20.275 ng/ml
9) t O-Xylene	13.929	665842	20.364 ng/ml
10) t 1,2,4-Trimethylbenzene	16.207	600347	21.819 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB110520\
Data File : FB025196.D
Signal(s) : FID2B.CH
Acq On : 5 Nov 2020 13:47
Operator : DD/AJ
Sample : L4621-05MS
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
PAV-B8-110320-10-20MS

Integration File: Calibration.e
Quant Time: Nov 05 15:20:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110320.M
Quant Title :
QLast Update : Wed Nov 04 08:05:53 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.689 min
Delta R.T.: 0.000 min
Response: 892958
Conc: 27.66 ng/ml

Instrument :
FID_B
ClientSampleId :
PAV-B8-110320-10-20MS




#2 2,2,4-Trimethylpentane

R.T.: 7.414 min
Delta R.T.: -0.004 min
Response: 881266
Conc: 26.56 ng/ml




#3 n-Heptane

R.T.: 7.755 min
Delta R.T.: -0.004 min
Response: 295914
Conc: 9.01 ng/ml



#4 Benzene


R.T.: 7.891 min
Delta R.T.: -0.002 min
Response: 389879
Conc: 9.81 ng/ml



#5 AAA-TFT


R.T.: 8.797 min
Delta R.T.: -0.002 min
Response: 445869
Conc: 19.75 ng/ml

Instrument :
FID_B
ClientSampleId :
PAV-B8-110320-10-20MS




#6 Toluene

R.T.: 10.629 min
Delta R.T.: -0.002 min
Response: 1096337
Conc: 29.32 ng/ml




#7 Ethylbenzene

R.T.: 13.068 min
Delta R.T.: -0.002 min
Response: 352991
Conc: 10.28 ng/ml



#8 m-Xylene


R.T.: 13.201 min
Delta R.T.: -0.002 min
Response: 684559
Conc: 20.27 ng/ml



#9 O-Xylene

R.T.: 13.929 min
Delta R.T.: -0.002 min
Response: 665842
Conc: 20.36 ng/ml

Instrument :
FID_B
ClientSampleId :
PAV-B8-110320-10-20MS



#10 1,2,4-Trimethylbenzene

R.T.: 16.207 min
Delta R.T.: -0.002 min
Response: 600347
Conc: 21.82 ng/ml