

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB110624\
 Data File : FBCCC05.D
 Signal(s) : FID2B.CH
 Acq On : 7 Nov 2024 2:29
 Operator : YP/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Nov 07 01:54:12 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB102824.M
 Quant Title :
 QLast Update : Mon Oct 28 12:19:57 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.793	585407	21.208 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.719	1277825	32.536 ng/ml
2) t 2,2,4-Trimethylpentane	7.425	1521855	32.493 ng/ml
3) t n-Heptane	7.755	469164	10.352 ng/ml
4) t Benzene	7.894	591578	11.067 ng/ml
6) t Toluene	10.622	1609184	32.224 ng/ml
7) t Ethylbenzene	13.058	485235	10.840 ng/ml
8) t m-Xylene	13.192	1083557	22.278 ng/ml
9) t O-Xylene	13.920	950466	20.851 ng/ml
10) t 1,2,4-Trimethylbenzene	16.195	875343	23.525 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB110624\
Data File : FBCCC05.D
Signal(s) : FID2B.CH
Acq On : 7 Nov 2024 2:29
Operator : YP/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Nov 07 01:54:12 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB102824.M
Quant Title :
QLast Update : Mon Oct 28 12:19:57 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.719 min
Delta R.T.: 0.000 min
Response: 1277825
Conc: 32.54 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.425 min
Delta R.T.: 0.000 min
Response: 1521855
Conc: 32.49 ng/ml




#3 n-Heptane

R.T.: 7.755 min
Delta R.T.: 0.002 min
Response: 469164
Conc: 10.35 ng/ml



#4 Benzene


R.T.: 7.894 min
Delta R.T.: 0.002 min
Response: 591578
Conc: 11.07 ng/ml



#5 AAA-TFT


R.T.: 8.793 min
Delta R.T.: 0.002 min
Response: 585407
Conc: 21.21 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.622 min
Delta R.T.: 0.003 min
Response: 1609184
Conc: 32.22 ng/ml




#7 Ethylbenzene

R.T.: 13.058 min
Delta R.T.: 0.003 min
Response: 485235
Conc: 10.84 ng/ml



#8 m-Xylene


R.T.: 13.192 min
Delta R.T.: 0.003 min
Response: 1083557
Conc: 22.28 ng/ml



#9 O-Xylene

R.T.: 13.920 min
Delta R.T.: 0.003 min
Response: 950466
Conc: 20.85 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.195 min
Delta R.T.: 0.003 min
Response: 875343
Conc: 23.53 ng/ml