

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB110918\
 Data File : FB018650.D
 Signal(s) : FID2B.CH
 Acq On : 10 Nov 2018 4:24
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Nov 10 08:39:31 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
 Quant Title :
 QLast Update : Fri Nov 09 10:53:13 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.109	172883	20.043 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.963	469039	32.069 ng/ml
2) t 2,2,4-Trimethylpentane	7.728	478230	32.602 ng/ml
3) t n-Heptane	8.065	151343	10.492 ng/ml
4) t Benzene	8.205	190883	10.535 ng/ml
6) t Toluene	10.959	484838	32.062 ng/ml
7) t Ethylbenzene	13.408	166757	10.693 ng/ml
8) t m-Xylene	13.540	353647	21.287 ng/ml
9) t O-Xylene	14.273	368468	21.477 ng/ml
10) t 1,2,4-Trimethylbenzene	16.556	428446	21.598 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB110918\
Data File : FB018650.D
Signal(s) : FID2B.CH
Acq On : 10 Nov 2018 4:24
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 23 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 10 08:39:31 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
Quant Title :
QLast Update : Fri Nov 09 10:53:13 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.963 min
Delta R.T.: 0.003 min
Response: 469039
Conc: 32.07 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.728 min
Delta R.T.: 0.005 min
Response: 478230
Conc: 32.60 ng/ml




#3 n-Heptane

R.T.: 8.065 min
Delta R.T.: 0.004 min
Response: 151343
Conc: 10.49 ng/ml




#4 Benzene

R.T.: 8.205 min
Delta R.T.: 0.005 min
Response: 190883
Conc: 10.54 ng/ml




#5 AAA-TFT

R.T.: 9.109 min
Delta R.T.: 0.004 min
Response: 172883
Conc: 20.04 ng/ml




#6 Toluene

R.T.: 10.959 min
Delta R.T.: 0.004 min
Response: 484838
Conc: 32.06 ng/ml




#7 Ethylbenzene

R.T.: 13.408 min
Delta R.T.: 0.005 min
Response: 166757
Conc: 10.69 ng/ml




#8 m-Xylene

R.T.: 13.540 min
Delta R.T.: 0.004 min
Response: 353647
Conc: 21.29 ng/ml



#9 O-Xylene

R.T.: 14.273 min
Delta R.T.: 0.005 min
Response: 368468
Conc: 21.48 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.556 min
Delta R.T.: 0.005 min
Response: 428446
Conc: 21.60 ng/ml