

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111120\
 Data File : FB025216.D
 Signal(s) : FID2B.CH
 Acq On : 11 Nov 2020 14:04
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: SAMPLE.e
 Quant Time: Nov 12 00:17:35 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110320.M
 Quant Title :
 QLast Update : Wed Nov 04 08:05:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.802	483738	21.424 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.692	1001414	31.025 ng/ml
2) t 2,2,4-Trimethylpentane	7.421	1009338	30.422 ng/ml
3) t n-Heptane	7.760	337618	10.279 ng/ml
4) t Benzene	7.896	420467	10.575 ng/ml
6) t Toluene	10.634	1176037	31.452 ng/ml
7) t Ethylbenzene	13.073	352971	10.279 ng/ml
8) t m-Xylene	13.207	713156	21.121 ng/ml
9) t O-Xylene	13.934	684225	20.926 ng/ml
10) t 1,2,4-Trimethylbenzene	16.211	596003	21.661 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111120\
Data File : FB025216.D
Signal(s) : FID2B.CH
Acq On : 11 Nov 2020 14:04
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: SAMPLE.e
Quant Time: Nov 12 00:17:35 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110320.M
Quant Title :
QLast Update : Wed Nov 04 08:05:53 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.692 min
Delta R.T.: 0.003 min
Response: 1001414
Conc: 31.02 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.421 min
Delta R.T.: 0.003 min
Response: 1009338
Conc: 30.42 ng/ml




#3 n-Heptane

R.T.: 7.760 min
Delta R.T.: 0.001 min
Response: 337618
Conc: 10.28 ng/ml




#4 Benzene

R.T.: 7.896 min
Delta R.T.: 0.003 min
Response: 420467
Conc: 10.58 ng/ml




#5 AAA-TFT

R.T.: 8.802 min
Delta R.T.: 0.003 min
Response: 483738
Conc: 21.42 ng/ml




#6 Toluene

R.T.: 10.634 min
Delta R.T.: 0.003 min
Response: 1176037
Conc: 31.45 ng/ml




#7 Ethylbenzene

R.T.: 13.073 min
Delta R.T.: 0.003 min
Response: 352971
Conc: 10.28 ng/ml




#8 m-Xylene

R.T.: 13.207 min
Delta R.T.: 0.003 min
Response: 713156
Conc: 21.12 ng/ml



#9 O-Xylene

R.T.: 13.934 min
Delta R.T.: 0.003 min
Response: 684225
Conc: 20.93 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.211 min
Delta R.T.: 0.003 min
Response: 596003
Conc: 21.66 ng/ml