

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111218\
 Data File : FB018705.D
 Signal(s) : FID2B.CH
 Acq On : 12 Nov 2018 10:38
 Operator : MJ/SJ
 Sample : BSF1112W2
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Nov 12 11:45:36 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
 Quant Title :
 QLast Update : Fri Nov 09 10:53:13 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.102	152143	17.638 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.956	401664	27.463 ng/ml
2) t	2,2,4-Trimethylpentane	7.721	403636	27.517 ng/ml
3) t	n-Heptane	8.060	128363	8.899 ng/ml
4) t	Benzene	8.198	159083	8.780 ng/ml
6) t	Toluene	10.952	427724	28.285 ng/ml
7) t	Ethylbenzene	13.402	149671	9.598 ng/ml
8) t	m-Xylene	13.535	317228	19.095 ng/ml
9) t	O-Xylene	14.267	321931	18.764 ng/ml
10) t	1,2,4-Trimethylbenzene	16.550	377893	19.050 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111218\
Data File : FB018705.D
Signal(s) : FID2B.CH
Acq On : 12 Nov 2018 10:38
Operator : MJ/SJ
Sample : BSF1112W2
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 12 11:45:36 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
Quant Title :
QLast Update : Fri Nov 09 10:53:13 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.956 min
Delta R.T.: -0.004 min
Response: 401664
Conc: 27.46 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.721 min
Delta R.T.: -0.002 min
Response: 403636
Conc: 27.52 ng/ml




#3 n-Heptane

R.T.: 8.060 min
Delta R.T.: -0.001 min
Response: 128363
Conc: 8.90 ng/ml




#4 Benzene

R.T.: 8.198 min
Delta R.T.: -0.002 min
Response: 159083
Conc: 8.78 ng/ml




#5 AAA-TFT

R.T.: 9.102 min
Delta R.T.: -0.003 min
Response: 152143
Conc: 17.64 ng/ml




#6 Toluene

R.T.: 10.952 min
Delta R.T.: -0.002 min
Response: 427724
Conc: 28.28 ng/ml




#7 Ethylbenzene

R.T.: 13.402 min
Delta R.T.: -0.001 min
Response: 149671
Conc: 9.60 ng/ml




#8 m-Xylene

R.T.: 13.535 min
Delta R.T.: -0.001 min
Response: 317228
Conc: 19.09 ng/ml



#9 O-Xylene

R.T.: 14.267 min
Delta R.T.: -0.001 min
Response: 321931
Conc: 18.76 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.550 min
Delta R.T.: -0.001 min
Response: 377893
Conc: 19.05 ng/ml