

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111218\
 Data File : FB018713.D
 Signal(s) : FID2B.CH
 Acq On : 12 Nov 2018 15:28
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Integration File: SAMPLE.e
 Quant Time: Nov 13 05:01:08 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
 Quant Title :
 QLast Update : Fri Nov 09 10:53:13 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.106	154127	17.868 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.960	464276	31.744 ng/ml
2) t 2,2,4-Trimethylpentane	7.724	469977	32.040 ng/ml
3) t n-Heptane	8.062	149704	10.378 ng/ml
4) t Benzene	8.200	190045	10.489 ng/ml
6) t Toluene	10.955	461752	30.535 ng/ml
7) t Ethylbenzene	13.405	156083	10.009 ng/ml
8) t m-Xylene	13.537	328336	19.763 ng/ml
9) t O-Xylene	14.269	344518	20.081 ng/ml
10) t 1,2,4-Trimethylbenzene	16.552	409901	20.663 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111218\
Data File : FB018713.D
Signal(s) : FID2B.CH
Acq On : 12 Nov 2018 15:28
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File: SAMPLE.e
Quant Time: Nov 13 05:01:08 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
Quant Title :
QLast Update : Fri Nov 09 10:53:13 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 464276
Conc: 31.74 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.724 min
Delta R.T.: 0.000 min
Response: 469977
Conc: 32.04 ng/ml




#3 n-Heptane

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 149704
Conc: 10.38 ng/ml




#4 Benzene

R.T.: 8.200 min
Delta R.T.: 0.000 min
Response: 190045
Conc: 10.49 ng/ml




#5 AAA-TFT

R.T.: 9.106 min
Delta R.T.: 0.001 min
Response: 154127
Conc: 17.87 ng/ml




#6 Toluene

R.T.: 10.955 min
Delta R.T.: 0.000 min
Response: 461752
Conc: 30.54 ng/ml




#7 Ethylbenzene

R.T.: 13.405 min
Delta R.T.: 0.001 min
Response: 156083
Conc: 10.01 ng/ml




#8 m-Xylene

R.T.: 13.537 min
Delta R.T.: 0.000 min
Response: 328336
Conc: 19.76 ng/ml



#9 O-Xylene

R.T.: 14.269 min
Delta R.T.: 0.002 min
Response: 344518
Conc: 20.08 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.552 min
Delta R.T.: 0.001 min
Response: 409901
Conc: 20.66 ng/ml