

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111418\
 Data File : FB018787.D
 Signal(s) : FID2B.CH
 Acq On : 14 Nov 2018 23:20
 Operator : MJ/SJ
 Sample : J5950-01 50X
 Misc : 5.00g/5.00ml DI WATER
 ALS Vial : 21 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Nov 15 04:50:26 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
 Quant Title :
 QLast Update : Fri Nov 09 10:53:13 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.170	497457	57.671 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.943	27665	1.892 ng/ml
2) t 2,2,4-Trimethylpentane	7.704	161858	11.034 ng/ml
3) t n-Heptane	8.062	240898	16.700 ng/ml
7) t Ethylbenzene	13.396	1349750	86.552 ng/ml
9) t O-Xylene	14.281	170379	9.931 ng/ml
10) t 1,2,4-Trimethylbenzene	16.552	5491698	276.840 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111418\
Data File : FB018787.D
Signal(s) : FID2B.CH
Acq On : 14 Nov 2018 23:20
Operator : MJ/SJ
Sample : J5950-01 50X
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 21 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 15 04:50:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
Quant Title :
QLast Update : Fri Nov 09 10:53:13 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.943 min
Delta R.T.: -0.017 min
Response: 27665
Conc: 1.89 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.704 min
Delta R.T.: -0.019 min
Response: 161858
Conc: 11.03 ng/ml




#3 n-Heptane

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 240898
Conc: 16.70 ng/ml




#5 AAA-TFT

R.T.: 9.170 min
Delta R.T.: 0.065 min
Response: 497457
Conc: 57.67 ng/ml




#7 Ethylbenzene

R.T.: 13.396 min
Delta R.T.: -0.007 min
Response: 1349750
Conc: 86.55 ng/ml



#9 O-Xylene

R.T.: 14.281 min
Delta R.T.: 0.013 min
Response: 170379
Conc: 9.93 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.552 min
Delta R.T.: 0.000 min
Response: 5491698
Conc: 276.84 ng/ml