

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111419\
 Data File : FB022988.D
 Signal(s) : FID2B.CH
 Acq On : 14 Nov 2019 14:40
 Operator : DD/AJ
 Sample : BSF1114W1
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Nov 15 00:44:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB103119.M
 Quant Title :
 QLast Update : Thu Oct 31 14:57:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	344211	15.290 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	697474	21.683 ng/ml
2) t 2,2,4-Trimethylpentane	7.415	809917	22.944 ng/ml
3) t n-Heptane	7.756	257282	7.396 ng/ml
4) t Benzene	7.890	348715	8.051 ng/ml
6) t Toluene	10.628	1021339	23.691 ng/ml
7) t Ethylbenzene	13.069	322169	8.070 ng/ml
8) t m-Xylene	13.202	617595	16.353 ng/ml
9) t O-Xylene	13.929	596587	16.739 ng/ml
10) t 1,2,4-Trimethylbenzene	16.208	499383	17.610 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111419\
Data File : FB022988.D
Signal(s) : FID2B.CH
Acq On : 14 Nov 2019 14:40
Operator : DD/AJ
Sample : BSF1114W1
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 15 00:44:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB103119.M
Quant Title :
QLast Update : Thu Oct 31 14:57:42 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.689 min
Delta R.T.: 0.000 min
Response: 697474
Conc: 21.68 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.415 min
Delta R.T.: -0.003 min
Response: 809917
Conc: 22.94 ng/ml




#3 n-Heptane

R.T.: 7.756 min
Delta R.T.: -0.003 min
Response: 257282
Conc: 7.40 ng/ml




#4 Benzene

R.T.: 7.890 min
Delta R.T.: -0.003 min
Response: 348715
Conc: 8.05 ng/ml




#5 AAA-TFT

R.T.: 8.797 min
Delta R.T.: -0.003 min
Response: 344211
Conc: 15.29 ng/ml




#6 Toluene

R.T.: 10.628 min
Delta R.T.: -0.003 min
Response: 1021339
Conc: 23.69 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min
Delta R.T.: -0.003 min
Response: 322169
Conc: 8.07 ng/ml




#8 m-Xylene

R.T.: 13.202 min
Delta R.T.: -0.003 min
Response: 617595
Conc: 16.35 ng/ml



#9 O-Xylene

R.T.: 13.929 min
Delta R.T.: -0.003 min
Response: 596587
Conc: 16.74 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min
Delta R.T.: -0.003 min
Response: 499383
Conc: 17.61 ng/ml