

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111424\
 Data File : FB031191.D
 Signal(s) : FID2B.CH
 Acq On : 15 Nov 2024 1:13
 Operator : YP/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Nov 15 03:11:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB102824.M
 Quant Title :
 QLast Update : Mon Oct 28 12:19:57 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.791	477640	17.304 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.718	1024536	26.087 ng/ml
2) t 2,2,4-Trimethylpentane	7.421	1272376	27.166 ng/ml
3) t n-Heptane	7.753	383098	8.453 ng/ml
4) t Benzene	7.891	485540	9.083 ng/ml
6) t Toluene	10.620	1321433	26.462 ng/ml
7) t Ethylbenzene	13.057	392950	8.779 ng/ml
8) t m-Xylene	13.190	821082	16.881 ng/ml
9) t O-Xylene	13.918	764160	16.764 ng/ml
10) t 1,2,4-Trimethylbenzene	16.194	617561	16.597 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111424\
Data File : FB031191.D
Signal(s) : FID2B.CH
Acq On : 15 Nov 2024 1:13
Operator : YP/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Nov 15 03:11:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB102824.M
Quant Title :
QLast Update : Mon Oct 28 12:19:57 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.718 min
Delta R.T.: -0.002 min
Response: 1024536
Conc: 26.09 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.421 min
Delta R.T.: -0.004 min
Response: 1272376
Conc: 27.17 ng/ml




#3 n-Heptane

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 383098
Conc: 8.45 ng/ml



#4 Benzene


R.T.: 7.891 min
Delta R.T.: 0.000 min
Response: 485540
Conc: 9.08 ng/ml



#5 AAA-TFT


R.T.: 8.791 min
Delta R.T.: 0.000 min
Response: 477640
Conc: 17.30 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.620 min
Delta R.T.: 0.000 min
Response: 1321433
Conc: 26.46 ng/ml




#7 Ethylbenzene

R.T.: 13.057 min
Delta R.T.: 0.002 min
Response: 392950
Conc: 8.78 ng/ml



#8 m-Xylene


R.T.: 13.190 min
Delta R.T.: 0.002 min
Response: 821082
Conc: 16.88 ng/ml



#9 O-Xylene

R.T.: 13.918 min
Delta R.T.: 0.001 min
Response: 764160
Conc: 16.76 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.194 min
Delta R.T.: 0.002 min
Response: 617561
Conc: 16.60 ng/ml