

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111518\
 Data File : FB018808.D
 Signal(s) : FID2B.CH
 Acq On : 15 Nov 2018 21:38
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc : 5.05g/5.00ml DI WATER
 ALS Vial : 18 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Nov 16 05:57:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
 Quant Title :
 QLast Update : Fri Nov 09 10:53:13 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.107	155325	18.007 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.959	403774	27.607 ng/ml
2) t	2,2,4-Trimethylpentane	7.722	435448	29.686 ng/ml
3) t	n-Heptane	8.063	134947	9.355 ng/ml
4) t	Benzene	8.201	173483	9.575 ng/ml
6) t	Toluene	10.956	447417	29.587 ng/ml
7) t	Ethylbenzene	13.405	156959	10.065 ng/ml
8) t	m-Xylene	13.538	331717	19.967 ng/ml
9) t	O-Xylene	14.270	343943	20.047 ng/ml
10) t	1,2,4-Trimethylbenzene	16.553	394414	19.883 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111518\
Data File : FB018808.D
Signal(s) : FID2B.CH
Acq On : 15 Nov 2018 21:38
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc : 5.05g/5.00ml DI WATER
ALS Vial : 18 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 16 05:57:16 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
Quant Title :
QLast Update : Fri Nov 09 10:53:13 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 403774
Conc: 27.61 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.722 min
Delta R.T.: 0.000 min
Response: 435448
Conc: 29.69 ng/ml




#3 n-Heptane

R.T.: 8.063 min
Delta R.T.: 0.002 min
Response: 134947
Conc: 9.36 ng/ml




#4 Benzene

R.T.: 8.201 min
Delta R.T.: 0.001 min
Response: 173483
Conc: 9.58 ng/ml




#5 AAA-TFT

R.T.: 9.107 min
Delta R.T.: 0.002 min
Response: 155325
Conc: 18.01 ng/ml




#6 Toluene

R.T.: 10.956 min
Delta R.T.: 0.002 min
Response: 447417
Conc: 29.59 ng/ml




#7 Ethylbenzene

R.T.: 13.405 min
Delta R.T.: 0.002 min
Response: 156959
Conc: 10.06 ng/ml




#8 m-Xylene

R.T.: 13.538 min
Delta R.T.: 0.002 min
Response: 331717
Conc: 19.97 ng/ml



#9 O-Xylene

R.T.: 14.270 min
Delta R.T.: 0.002 min
Response: 343943
Conc: 20.05 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.553 min
Delta R.T.: 0.003 min
Response: 394414
Conc: 19.88 ng/ml