

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111819\
Data File : FB023022.D
Signal(s) : FID2B.CH
Acq On : 18 Nov 2019 14:16
Operator : DD/AJ
Sample : K5669-02
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 4 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 19 00:39:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB103119.M
Quant Title :
QLast Update : Thu Oct 31 14:57:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.796	447562	19.881 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.690	2466	0.077 ng/ml
2) t 2,2,4-Trimethylpentane	7.385	4696	0.133 ng/ml
4) t Benzene	7.915	996	0.023 ng/ml
6) t Toluene	10.635	4397	0.102 ng/ml
8) t m-Xylene	13.199	3294	0.087 ng/ml
9) t O-Xylene	13.946	4520	0.127 ng/ml
10) t 1,2,4-Trimethylbenzene	16.213	2015	0.071 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111819\
Data File : FB023022.D
Signal(s) : FID2B.CH
Acq On : 18 Nov 2019 14:16
Operator : DD/AJ
Sample : K5669-02
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 4 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 19 00:39:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB103119.M
Quant Title :
QLast Update : Thu Oct 31 14:57:42 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.690 min
Delta R.T.: 0.000 min
Response: 2466
Conc: 0.08 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.385 min
Delta R.T.: -0.033 min
Response: 4696
Conc: 0.13 ng/ml




#4 Benzene

R.T.: 7.915 min
Delta R.T.: 0.022 min
Response: 996
Conc: 0.02 ng/ml




#5 AAA-TFT

R.T.: 8.796 min
Delta R.T.: -0.004 min
Response: 447562
Conc: 19.88 ng/ml




#6 Toluene

R.T.: 10.635 min
Delta R.T.: 0.004 min
Response: 4397
Conc: 0.10 ng/ml




#8 m-Xylene

R.T.: 13.199 min
Delta R.T.: -0.006 min
Response: 3294
Conc: 0.09 ng/ml



#9 o-Xylene

R.T.: 13.946 min
Delta R.T.: 0.013 min
Response: 4520
Conc: 0.13 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.213 min
Delta R.T.: 0.001 min
Response: 2015
Conc: 0.07 ng/ml