

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111819\
 Data File : FB023026.D
 Signal(s) : FID2B.CH
 Acq On : 18 Nov 2019 16:18
 Operator : DD/AJ
 Sample : K5676-17
 Misc : 5.00g/5.00ml DI WATER
 ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Nov 19 00:39:49 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB103119.M
 Quant Title :
 QLast Update : Thu Oct 31 14:57:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.800	368551	16.371 ng/ml
Target Compounds			
2) t 2,2,4-Trimethylpentane	7.415	142	0.004 ng/ml
4) t Benzene	7.900	1014	0.023 ng/ml
6) t Toluene	10.632	9570	0.222 ng/ml
7) t Ethylbenzene	13.058	70760	1.772 ng/ml
8) t m-Xylene	13.202	84639	2.241 ng/ml
9) t O-Xylene	13.934	105232	2.953 ng/ml
10) t 1,2,4-Trimethylbenzene	16.212	241193	8.505 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111819\
Data File : FB023026.D
Signal(s) : FID2B.CH
Acq On : 18 Nov 2019 16:18
Operator : DD/AJ
Sample : K5676-17
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 19 00:39:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB103119.M
Quant Title :
QLast Update : Thu Oct 31 14:57:42 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#2 2,2,4-Trimethylpentane

R.T.: 7.415 min
Delta R.T.: -0.003 min
Response: 142
Conc: 0.00 ng/ml




#4 Benzene

R.T.: 7.900 min
Delta R.T.: 0.007 min
Response: 1014
Conc: 0.02 ng/ml




#5 AAA-TFT

R.T.: 8.800 min
Delta R.T.: 0.000 min
Response: 368551
Conc: 16.37 ng/ml




#6 Toluene

R.T.: 10.632 min
Delta R.T.: 0.000 min
Response: 9570
Conc: 0.22 ng/ml




#7 Ethylbenzene

R.T.: 13.058 min
Delta R.T.: -0.014 min
Response: 70760
Conc: 1.77 ng/ml




#8 m-Xylene

R.T.: 13.202 min
Delta R.T.: -0.003 min
Response: 84639
Conc: 2.24 ng/ml



#9 o-Xylene

R.T.: 13.934 min
Delta R.T.: 0.001 min
Response: 105232
Conc: 2.95 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.212 min
Delta R.T.: 0.001 min
Response: 241193
Conc: 8.51 ng/ml