

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112124\
 Data File : FB031254.D
 Signal(s) : FID2B.CH
 Acq On : 21 Nov 2024 14:05
 Operator : YP/AJ
 Sample : FB112124GROICV
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 FB112124GROICV

Integration File: Calibration.e
 Quant Time: Nov 21 13:14:45 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112124.M
 Quant Title :
 QLast Update : Thu Nov 21 12:34:29 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	603312	21.434 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.721	1121886	28.447 ng/ml
2) t 2,2,4-Trimethylpentane	7.428	1306226	27.623 ng/ml
3) t n-Heptane	7.758	435889	9.333 ng/ml
4) t Benzene	7.896	509347	9.398 ng/ml
6) t Toluene	10.626	1373087	27.457 ng/ml
7) t Ethylbenzene	13.063	396539	9.039 ng/ml
8) t m-Xylene	13.197	853865	18.111 ng/ml
9) t O-Xylene	13.925	797956	18.321 ng/ml
10) t 1,2,4-Trimethylbenzene	16.201	644383	18.052 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112124\
Data File : FB031254.D
Signal(s) : FID2B.CH
Acq On : 21 Nov 2024 14:05
Operator : YP/AJ
Sample : FB112124GROICV
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
FB112124GROICV

Integration File: Calibration.e
Quant Time: Nov 21 13:14:45 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112124.M
Quant Title :
QLast Update : Thu Nov 21 12:34:29 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.721 min
Delta R.T.: -0.001 min
Response: 1121886
Conc: 28.45 ng/ml

Instrument :
FID_B
ClientSampleId :
FB112124GROICV




#2 2,2,4-Trimethylpentane

R.T.: 7.428 min
Delta R.T.: 0.000 min
Response: 1306226
Conc: 27.62 ng/ml




#3 n-Heptane

R.T.: 7.758 min
Delta R.T.: 0.000 min
Response: 435889
Conc: 9.33 ng/ml



#4 Benzene


R.T.: 7.896 min
Delta R.T.: 0.000 min
Response: 509347
Conc: 9.40 ng/ml



#5 AAA-TFT


R.T.: 8.797 min
Delta R.T.: 0.000 min
Response: 603312
Conc: 21.43 ng/ml

Instrument :
FID_B
ClientSampleId :
FB112124GROICV




#6 Toluene

R.T.: 10.626 min
Delta R.T.: 0.002 min
Response: 1373087
Conc: 27.46 ng/ml




#7 Ethylbenzene

R.T.: 13.063 min
Delta R.T.: 0.002 min
Response: 396539
Conc: 9.04 ng/ml



#8 m-Xylene


R.T.: 13.197 min
Delta R.T.: 0.002 min
Response: 853865
Conc: 18.11 ng/ml



#9 O-Xylene

R.T.: 13.925 min
Delta R.T.: 0.002 min
Response: 797956
Conc: 18.32 ng/ml

Instrument :
FID_B
ClientSampleId :
FB112124GROICV



#10 1,2,4-Trimethylbenzene

R.T.: 16.201 min
Delta R.T.: 0.002 min
Response: 644383
Conc: 18.05 ng/ml