

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112219\
 Data File : FB023062.D
 Signal(s) : FID2B.CH
 Acq On : 22 Nov 2019 13:20
 Operator : DD/AJ
 Sample : FB102219GROICV
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Nov 22 12:44:08 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
 Quant Title :
 QLast Update : Fri Nov 22 12:34:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.799	490531	19.970 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.690	990157	30.300 ng/ml
2) t 2,2,4-Trimethylpentane	7.418	1055607	30.272 ng/ml
3) t n-Heptane	7.758	346435	10.076 ng/ml
4) t Benzene	7.892	421401	10.114 ng/ml
6) t Toluene	10.630	1229640	30.524 ng/ml
7) t Ethylbenzene	13.071	380866	9.870 ng/ml
8) t m-Xylene	13.204	707943	19.454 ng/ml
9) t O-Xylene	13.932	680268	19.593 ng/ml
10) t 1,2,4-Trimethylbenzene	16.211	569533	20.309 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112219\
Data File : FB023062.D
Signal(s) : FID2B.CH
Acq On : 22 Nov 2019 13:20
Operator : DD/AJ
Sample : FB102219GR0ICV
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Nov 22 12:44:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.690 min
Delta R.T.: 0.000 min
Response: 990157
Conc: 30.30 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 1055607
Conc: 30.27 ng/ml




#3 n-Heptane

R.T.: 7.758 min
Delta R.T.: 0.000 min
Response: 346435
Conc: 10.08 ng/ml



#4 Benzene


R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 421401
Conc: 10.11 ng/ml



#5 AAA-TFT


R.T.: 8.799 min
Delta R.T.: 0.000 min
Response: 490531
Conc: 19.97 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.000 min
Response: 1229640
Conc: 30.52 ng/ml




#7 Ethylbenzene

R.T.: 13.071 min
Delta R.T.: 0.000 min
Response: 380866
Conc: 9.87 ng/ml



#8 m-Xylene


R.T.: 13.204 min
Delta R.T.: 0.000 min
Response: 707943
Conc: 19.45 ng/ml



#9 O-Xylene

R.T.: 13.932 min
Delta R.T.: 0.000 min
Response: 680268
Conc: 19.59 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.211 min
Delta R.T.: 0.000 min
Response: 569533
Conc: 20.31 ng/ml