

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112219\
 Data File : FB023065.D
 Signal(s) : FID2B.CH
 Acq On : 22 Nov 2019 15:06
 Operator : DD/AJ
 Sample : BSF1122S1
 Misc : 5.02g/5.00ml DI WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF1122S1

Integration File: Calibration.e
 Quant Time: Nov 22 20:47:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
 Quant Title :
 QLast Update : Fri Nov 22 12:34:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.799	493062	20.073 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.691	960363	29.388 ng/ml
2) t 2,2,4-Trimethylpentane	7.418	1059585	30.386 ng/ml
3) t n-Heptane	7.758	347389	10.103 ng/ml
4) t Benzene	7.892	399748	9.594 ng/ml
6) t Toluene	10.631	1189451	29.527 ng/ml
7) t Ethylbenzene	13.071	371488	9.627 ng/ml
8) t m-Xylene	13.204	700027	19.237 ng/ml
9) t O-Xylene	13.932	672457	19.368 ng/ml
10) t 1,2,4-Trimethylbenzene	16.211	552366	19.697 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112219\
Data File : FB023065.D
Signal(s) : FID2B.CH
Acq On : 22 Nov 2019 15:06
Operator : DD/AJ
Sample : BSF1122S1
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF1122S1

Integration File: Calibration.e
Quant Time: Nov 22 20:47:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.691 min
Delta R.T.: 0.000 min
Response: 960363
Conc: 29.39 ng/ml

Instrument :
FID_B
ClientSampleId :
BSF1122S1




#2 2,2,4-Trimethylpentane

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 1059585
Conc: 30.39 ng/ml




#3 n-Heptane

R.T.: 7.758 min
Delta R.T.: 0.000 min
Response: 347389
Conc: 10.10 ng/ml



#4 Benzene


R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 399748
Conc: 9.59 ng/ml



#5 AAA-TFT


R.T.: 8.799 min
Delta R.T.: 0.000 min
Response: 493062
Conc: 20.07 ng/ml

Instrument :
FID_B
ClientSampleId :
BSF1122S1




#6 Toluene

R.T.: 10.631 min
Delta R.T.: 0.000 min
Response: 1189451
Conc: 29.53 ng/ml




#7 Ethylbenzene

R.T.: 13.071 min
Delta R.T.: 0.000 min
Response: 371488
Conc: 9.63 ng/ml



#8 m-Xylene


R.T.: 13.204 min
Delta R.T.: 0.000 min
Response: 700027
Conc: 19.24 ng/ml



#9 O-Xylene

R.T.: 13.932 min
Delta R.T.: 0.000 min
Response: 672457
Conc: 19.37 ng/ml

Instrument :
FID_B
ClientSampleId :
BSF1122S1



#10 1,2,4-Trimethylbenzene

R.T.: 16.211 min
Delta R.T.: 0.000 min
Response: 552366
Conc: 19.70 ng/ml