

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112519\
 Data File : FB023083.D
 Signal(s) : FID2B.CH
 Acq On : 25 Nov 2019 11:06
 Operator : DD/AJ
 Sample : VBF1125S1
 Misc : 5.00g/5.00ml DI WATER
 ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Nov 25 12:00:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
 Quant Title :
 QLast Update : Fri Nov 22 12:34:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.795	500155	20.361 ng/ml
Target Compounds			
4) t Benzene	7.887	53	0.001 ng/ml
6) t Toluene	10.623	6618	0.164 ng/ml
8) t m-Xylene	13.209	3505	0.096 ng/ml
9) t O-Xylene	13.928	850	0.024 ng/ml
10) t 1,2,4-Trimethylbenzene	16.198	3970	0.142 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112519\
Data File : FB023083.D
Signal(s) : FID2B.CH
Acq On : 25 Nov 2019 11:06
Operator : DD/AJ
Sample : VBF1125S1
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 25 12:00:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#4 Benzene

R.T.: 7.887 min
Delta R.T.: -0.005 min
Response: 53
Conc: 0.00 ng/ml




#5 AAA-TFT

R.T.: 8.795 min
Delta R.T.: -0.004 min
Response: 500155
Conc: 20.36 ng/ml




#6 Toluene

R.T.: 10.623 min
Delta R.T.: -0.008 min
Response: 6618
Conc: 0.16 ng/ml




#8 m-Xylene

R.T.: 13.209 min
Delta R.T.: 0.005 min
Response: 3505
Conc: 0.10 ng/ml



#9 O-Xylene

R.T.: 13.928 min
Delta R.T.: -0.004 min
Response: 850
Conc: 0.02 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.198 min
Delta R.T.: -0.012 min
Response: 3970
Conc: 0.14 ng/ml