

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112519\
 Data File : FB023089.D
 Signal(s) : FID2B.CH
 Acq On : 25 Nov 2019 17:04
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File: SAMPLE.e
 Quant Time: Nov 26 00:26:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
 Quant Title :
 QLast Update : Fri Nov 22 12:34:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.795	522169	21.258 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	1013914	31.027 ng/ml
2) t 2,2,4-Trimethylpentane	7.415	1121168	32.152 ng/ml
3) t n-Heptane	7.754	366952	10.672 ng/ml
4) t Benzene	7.889	438475	10.524 ng/ml
6) t Toluene	10.628	1251808	31.075 ng/ml
7) t Ethylbenzene	13.069	389828	10.102 ng/ml
8) t m-Xylene	13.201	735903	20.223 ng/ml
9) t O-Xylene	13.929	702708	20.239 ng/ml
10) t 1,2,4-Trimethylbenzene	16.208	578408	20.625 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112519\
Data File : FB023089.D
Signal(s) : FID2B.CH
Acq On : 25 Nov 2019 17:04
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File: SAMPLE.e
Quant Time: Nov 26 00:26:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.688 min
Delta R.T.: -0.003 min
Response: 1013914
Conc: 31.03 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.415 min
Delta R.T.: -0.003 min
Response: 1121168
Conc: 32.15 ng/ml




#3 n-Heptane

R.T.: 7.754 min
Delta R.T.: -0.004 min
Response: 366952
Conc: 10.67 ng/ml




#4 Benzene

R.T.: 7.889 min
Delta R.T.: -0.004 min
Response: 438475
Conc: 10.52 ng/ml




#5 AAA-TFT

R.T.: 8.795 min
Delta R.T.: -0.004 min
Response: 522169
Conc: 21.26 ng/ml




#6 Toluene

R.T.: 10.628 min
Delta R.T.: -0.003 min
Response: 1251808
Conc: 31.07 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min
Delta R.T.: -0.002 min
Response: 389828
Conc: 10.10 ng/ml




#8 m-Xylene

R.T.: 13.201 min
Delta R.T.: -0.002 min
Response: 735903
Conc: 20.22 ng/ml



#9 O-Xylene

R.T.: 13.929 min
Delta R.T.: -0.002 min
Response: 702708
Conc: 20.24 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min
Delta R.T.: -0.002 min
Response: 578408
Conc: 20.63 ng/ml