

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112520\
 Data File : FB025279.D
 Signal(s) : FID2B.CH
 Acq On : 25 Nov 2020 11:59
 Operator : DD/AJ
 Sample : L4868-04MS
 Misc : 5.00g/5.00ml DI WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 KENS-B3-112020-8-10MS

Integration File: Calibration.e
 Quant Time: Nov 26 00:17:14 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110320.M
 Quant Title :
 QLast Update : Wed Nov 04 08:05:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.796	476438	21.100 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.685	879701	27.254 ng/ml
2) t 2,2,4-Trimethylpentane	7.413	783474	23.614 ng/ml
3) t n-Heptane	7.754	259205	7.892 ng/ml
4) t Benzene	7.889	417962	10.512 ng/ml
6) t Toluene	10.628	1156956	30.942 ng/ml
7) t Ethylbenzene	13.069	342075	9.961 ng/ml
8) t m-Xylene	13.202	684589	20.275 ng/ml
9) t O-Xylene	13.929	667350	20.410 ng/ml
10) t 1,2,4-Trimethylbenzene	16.207	553552	20.118 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112520\
Data File : FB025279.D
Signal(s) : FID2B.CH
Acq On : 25 Nov 2020 11:59
Operator : DD/AJ
Sample : L4868-04MS
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
KENS-B3-112020-8-10MS

Integration File: Calibration.e
Quant Time: Nov 26 00:17:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110320.M
Quant Title :
QLast Update : Wed Nov 04 08:05:53 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.685 min
Delta R.T.: -0.004 min
Response: 879701
Conc: 27.25 ng/ml

Instrument :
FID_B
ClientSampleId :
KENS-B3-112020-8-10MS




#2 2,2,4-Trimethylpentane

R.T.: 7.413 min
Delta R.T.: -0.005 min
Response: 783474
Conc: 23.61 ng/ml




#3 n-Heptane

R.T.: 7.754 min
Delta R.T.: -0.005 min
Response: 259205
Conc: 7.89 ng/ml



#4 Benzene


R.T.: 7.889 min
Delta R.T.: -0.004 min
Response: 417962
Conc: 10.51 ng/ml



#5 AAA-TFT


R.T.: 8.796 min
Delta R.T.: -0.003 min
Response: 476438
Conc: 21.10 ng/ml

Instrument :
FID_B
ClientSampleId :
KENS-B3-112020-8-10MS




#6 Toluene

R.T.: 10.628 min
Delta R.T.: -0.002 min
Response: 1156956
Conc: 30.94 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min
Delta R.T.: -0.002 min
Response: 342075
Conc: 9.96 ng/ml



#8 m-Xylene


R.T.: 13.202 min
Delta R.T.: -0.002 min
Response: 684589
Conc: 20.28 ng/ml



#9 O-Xylene

R.T.: 13.929 min
Delta R.T.: -0.002 min
Response: 667350
Conc: 20.41 ng/ml

Instrument :
FID_B
ClientSampleId :
KENS-B3-112020-8-10MS



#10 1,2,4-Trimethylbenzene

R.T.: 16.207 min
Delta R.T.: -0.002 min
Response: 553552
Conc: 20.12 ng/ml