

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112618\
Data File : FB018870.D
Signal(s) : FID2B.CH
Acq On : 26 Nov 2018 18:56
Operator : MJ/SJ
Sample : J6065-01 50X
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 27 03:38:54 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
Quant Title :
QLast Update : Fri Nov 09 10:53:13 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.114	179168	20.771 ng/ml
Target Compounds				
3) t	n-Heptane	8.053	6184	0.429 ng/ml
4) t	Benzene	8.213	7426	0.410 ng/ml
6) t	Toluene	10.965	21877	1.447 ng/ml
7) t	Ethylbenzene	13.406	29980	1.922 ng/ml
8) t	m-Xylene	13.544	42754	2.573 ng/ml
9) t	O-Xylene	14.285	21642	1.261 ng/ml
10) t	1,2,4-Trimethylbenzene	16.566	35677	1.799 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112618\
Data File : FB018870.D
Signal(s) : FID2B.CH
Acq On : 26 Nov 2018 18:56
Operator : MJ/SJ
Sample : J6065-01 50X
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 27 03:38:54 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
Quant Title :
QLast Update : Fri Nov 09 10:53:13 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#3 n-Heptane

R.T.: 8.053 min
Delta R.T.: -0.008 min
Response: 6184
Conc: 0.43 ng/ml




#4 Benzene

R.T.: 8.213 min
Delta R.T.: 0.014 min
Response: 7426
Conc: 0.41 ng/ml




#5 AAA-TFT

R.T.: 9.114 min
Delta R.T.: 0.009 min
Response: 179168
Conc: 20.77 ng/ml




#6 Toluene

R.T.: 10.965 min
Delta R.T.: 0.011 min
Response: 21877
Conc: 1.45 ng/ml




#7 Ethylbenzene

R.T.: 13.406 min
Delta R.T.: 0.003 min
Response: 29980
Conc: 1.92 ng/ml




#8 m-Xylene

R.T.: 13.544 min
Delta R.T.: 0.008 min
Response: 42754
Conc: 2.57 ng/ml



#9 O-Xylene

R.T.: 14.285 min
Delta R.T.: 0.017 min
Response: 21642
Conc: 1.26 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.566 min
Delta R.T.: 0.015 min
Response: 35677
Conc: 1.80 ng/ml