

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112619\
Data File : FB023104.D
Signal(s) : FID2B.CH
Acq On : 26 Nov 2019 18:16
Operator : DD/AJ
Sample : K5667-02
Misc : 5.04g/5.00ml DI WATER
ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 27 00:13:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.796	409150	16.657 ng/ml
Target Compounds			
6) t Toluene	10.630	5852	0.145 ng/ml
7) t Ethylbenzene	13.084	1641	0.043 ng/ml
8) t m-Xylene	13.202	4781	0.131 ng/ml
9) t O-Xylene	13.924	3735	0.108 ng/ml
10) t 1,2,4-Trimethylbenzene	16.206	9349	0.333 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112619\
Data File : FB023104.D
Signal(s) : FID2B.CH
Acq On : 26 Nov 2019 18:16
Operator : DD/AJ
Sample : K5667-02
Misc : 5.04g/5.00ml DI WATER
ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 27 00:13:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT

R.T.: 8.796 min
Delta R.T.: -0.003 min
Response: 409150
Conc: 16.66 ng/ml




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.000 min
Response: 5852
Conc: 0.15 ng/ml




#7 Ethylbenzene

R.T.: 13.084 min
Delta R.T.: 0.013 min
Response: 1641
Conc: 0.04 ng/ml




#8 m-Xylene

R.T.: 13.202 min
Delta R.T.: -0.002 min
Response: 4781
Conc: 0.13 ng/ml



#9 O-Xylene

R.T.: 13.924 min
Delta R.T.: -0.008 min
Response: 3735
Conc: 0.11 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.206 min
Delta R.T.: -0.004 min
Response: 9349
Conc: 0.33 ng/ml