

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID\_B\Data\FB112718\  
 Data File : FB018874.D  
 Signal(s) : FID2B.CH  
 Acq On : 27 Nov 2018 9:23  
 Operator : MJ/SJ  
 Sample : BSF1127S1  
 Misc : 5.00g/5.00ml DI WATER  
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e  
 Quant Time: Nov 27 14:39:46 2018  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB110918.M  
 Quant Title :  
 QLast Update : Fri Nov 09 10:53:13 2018  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 5 g/ml  
 Signal Phase : RTX-502.2  
 Signal Info : 60mx0.53mmx3.00um

| Compound                     | R.T.   | Response | Conc Units   |
|------------------------------|--------|----------|--------------|
| -----                        |        |          |              |
| System Monitoring Compounds  |        |          |              |
| 5) s AAA-TFT                 | 9.111  | 175425   | 20.337 ng/ml |
| Target Compounds             |        |          |              |
| 1) t 2-Methylpentane         | 4.963  | 448136   | 30.640 ng/ml |
| 2) t 2,2,4-Trimethylpentane  | 7.727  | 460537   | 31.396 ng/ml |
| 3) t n-Heptane               | 8.067  | 149479   | 10.363 ng/ml |
| 4) t Benzene                 | 8.205  | 189540   | 10.461 ng/ml |
| 6) t Toluene                 | 10.960 | 476078   | 31.483 ng/ml |
| 7) t Ethylbenzene            | 13.411 | 160519   | 10.293 ng/ml |
| 8) t m-Xylene                | 13.544 | 343168   | 20.656 ng/ml |
| 9) t O-Xylene                | 14.276 | 353704   | 20.616 ng/ml |
| 10) t 1,2,4-Trimethylbenzene | 16.560 | 424839   | 21.416 ng/ml |
| -----                        |        |          |              |


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID\_B\Data\FB112718\  
Data File : FB018874.D  
Signal(s) : FID2B.CH  
Acq On : 27 Nov 2018 9:23  
Operator : MJ/SJ  
Sample : BSF1127S1  
Misc : 5.00g/5.00ml DI WATER  
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e  
Quant Time: Nov 27 14:39:46 2018  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB110918.M  
Quant Title :  
QLast Update : Fri Nov 09 10:53:13 2018  
Response via : Initial Calibration  
Integrator: ChemStation


Volume Inj. : 5 g/ml  
Signal Phase : RTX-502.2  
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.963 min  
Delta R.T.: 0.003 min  
Response: 448136  
Conc: 30.64 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.727 min  
Delta R.T.: 0.003 min  
Response: 460537  
Conc: 31.40 ng/ml




#3 n-Heptane

R.T.: 8.067 min  
Delta R.T.: 0.006 min  
Response: 149479  
Conc: 10.36 ng/ml




#4 Benzene

R.T.: 8.205 min  
Delta R.T.: 0.006 min  
Response: 189540  
Conc: 10.46 ng/ml




#5 AAA-TFT

R.T.: 9.111 min  
Delta R.T.: 0.006 min  
Response: 175425  
Conc: 20.34 ng/ml




#6 Toluene

R.T.: 10.960 min  
Delta R.T.: 0.006 min  
Response: 476078  
Conc: 31.48 ng/ml




#7 Ethylbenzene

R.T.: 13.411 min  
Delta R.T.: 0.008 min  
Response: 160519  
Conc: 10.29 ng/ml




#8 m-Xylene

R.T.: 13.544 min  
Delta R.T.: 0.008 min  
Response: 343168  
Conc: 20.66 ng/ml



#9 O-Xylene

R.T.: 14.276 min  
Delta R.T.: 0.008 min  
Response: 353704  
Conc: 20.62 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.560 min  
Delta R.T.: 0.009 min  
Response: 424839  
Conc: 21.42 ng/ml