

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112822\
 Data File : FB029057.D
 Signal(s) : FID2B.CH
 Acq On : 28 Nov 2022 10:35
 Operator : AJ/MA
 Sample : 20 GRO STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Nov 28 10:47:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112822.M
 Quant Title :
 QLast Update : Mon Nov 28 10:47:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.791	228953	20.000 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.713	334973	30.000 ng/ml
2) t 2,2,4-Trimethylpentane	7.417	412490	30.000 ng/ml
3) t n-Heptane	7.751	106874	10.000 ng/ml
4) t Benzene	7.888	151533	10.000 ng/ml
6) t Toluene	10.620	455592	30.000 ng/ml
7) t Ethylbenzene	13.056	150661	10.000 ng/ml
8) t m-Xylene	13.189	302525	20.000 ng/ml
9) t O-Xylene	13.916	308098	20.000 ng/ml
10) t 1,2,4-Trimethylbenzene	16.191	333368	20.000 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB112822\
Data File : FB029057.D
Signal(s) : FID2B.CH
Acq On : 28 Nov 2022 10:35
Operator : AJ/MA
Sample : 20 GRO STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Nov 28 10:47:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112822.M
Quant Title :
QLast Update : Mon Nov 28 10:47:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.713 min
Delta R.T.: 0.000 min
Response: 334973
Conc: 30.00 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.417 min
Delta R.T.: 0.000 min
Response: 412490
Conc: 30.00 ng/ml




#3 n-Heptane

R.T.: 7.751 min
Delta R.T.: 0.000 min
Response: 106874
Conc: 10.00 ng/ml



#4 Benzene


R.T.: 7.888 min
Delta R.T.: 0.000 min
Response: 151533
Conc: 10.00 ng/ml



#5 AAA-TFT


R.T.: 8.791 min
Delta R.T.: 0.000 min
Response: 228953
Conc: 20.00 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.620 min
Delta R.T.: 0.000 min
Response: 455592
Conc: 30.00 ng/ml




#7 Ethylbenzene

R.T.: 13.056 min
Delta R.T.: 0.000 min
Response: 150661
Conc: 10.00 ng/ml



#8 m-Xylene


R.T.: 13.189 min
Delta R.T.: 0.000 min
Response: 302525
Conc: 20.00 ng/ml



#9 O-Xylene

R.T.: 13.916 min
Delta R.T.: 0.000 min
Response: 308098
Conc: 20.00 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.191 min
Delta R.T.: 0.000 min
Response: 333368
Conc: 20.00 ng/ml