

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID_B\Data\FB113018\
 Data File : FB018905.D
 Signal(s) : FID2B.CH
 Acq On : 30 Nov 2018 15:51
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Nov 30 16:15:17 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
 Quant Title :
 QLast Update : Fri Nov 09 10:53:13 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.109	150031	17.393 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.962	387766	26.513 ng/ml
2) t 2,2,4-Trimethylpentane	7.728	397376	27.090 ng/ml
3) t n-Heptane	8.065	128011	8.874 ng/ml
4) t Benzene	8.205	162867	8.989 ng/ml
6) t Toluene	10.959	422187	27.919 ng/ml
7) t Ethylbenzene	13.408	145800	9.349 ng/ml
8) t m-Xylene	13.541	309259	18.615 ng/ml
9) t O-Xylene	14.273	325662	18.982 ng/ml
10) t 1,2,4-Trimethylbenzene	16.556	385580	19.437 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID_B\Data\FB113018\
Data File : FB018905.D
Signal(s) : FID2B.CH
Acq On : 30 Nov 2018 15:51
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Nov 30 16:15:17 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB110918.M
Quant Title :
QLast Update : Fri Nov 09 10:53:13 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.962 min
Delta R.T.: 0.002 min
Response: 387766
Conc: 26.51 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.728 min
Delta R.T.: 0.005 min
Response: 397376
Conc: 27.09 ng/ml




#3 n-Heptane

R.T.: 8.065 min
Delta R.T.: 0.004 min
Response: 128011
Conc: 8.87 ng/ml




#4 Benzene

R.T.: 8.205 min
Delta R.T.: 0.005 min
Response: 162867
Conc: 8.99 ng/ml




#5 AAA-TFT

R.T.: 9.109 min
Delta R.T.: 0.004 min
Response: 150031
Conc: 17.39 ng/ml




#6 Toluene

R.T.: 10.959 min
Delta R.T.: 0.005 min
Response: 422187
Conc: 27.92 ng/ml




#7 Ethylbenzene

R.T.: 13.408 min
Delta R.T.: 0.005 min
Response: 145800
Conc: 9.35 ng/ml




#8 m-Xylene

R.T.: 13.541 min
Delta R.T.: 0.005 min
Response: 309259
Conc: 18.62 ng/ml



#9 O-Xylene

R.T.: 14.273 min
Delta R.T.: 0.005 min
Response: 325662
Conc: 18.98 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.556 min
Delta R.T.: 0.006 min
Response: 385580
Conc: 19.44 ng/ml