

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB120219\
 Data File : FB023126.D
 Signal(s) : FID2B.CH
 Acq On : 2 Dec 2019 16:10
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Dec 02 15:32:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
 Quant Title :
 QLast Update : Fri Nov 22 12:34:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	462959	18.847 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	881948	26.989 ng/ml
2) t 2,2,4-Trimethylpentane	7.416	975050	27.962 ng/ml
3) t n-Heptane	7.756	312540	9.090 ng/ml
4) t Benzene	7.890	385520	9.253 ng/ml
6) t Toluene	10.630	1137071	28.227 ng/ml
7) t Ethylbenzene	13.071	344140	8.918 ng/ml
8) t m-Xylene	13.204	651503	17.903 ng/ml
9) t O-Xylene	13.932	632414	18.214 ng/ml
10) t 1,2,4-Trimethylbenzene	16.211	527283	18.802 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB120219\
Data File : FB023126.D
Signal(s) : FID2B.CH
Acq On : 2 Dec 2019 16:10
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Dec 02 15:32:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.688 min
Delta R.T.: -0.003 min
Response: 881948
Conc: 26.99 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min
Delta R.T.: -0.002 min
Response: 975050
Conc: 27.96 ng/ml




#3 n-Heptane

R.T.: 7.756 min
Delta R.T.: -0.001 min
Response: 312540
Conc: 9.09 ng/ml



#4 Benzene


R.T.: 7.890 min
Delta R.T.: -0.002 min
Response: 385520
Conc: 9.25 ng/ml



#5 AAA-TFT


R.T.: 8.797 min
Delta R.T.: -0.002 min
Response: 462959
Conc: 18.85 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.000 min
Response: 1137071
Conc: 28.23 ng/ml




#7 Ethylbenzene

R.T.: 13.071 min
Delta R.T.: 0.000 min
Response: 344140
Conc: 8.92 ng/ml



#8 m-Xylene

R.T.: 13.204 min
Delta R.T.: 0.000 min
Response: 651503
Conc: 17.90 ng/ml



#9 O-Xylene


R.T.: 13.932 min

Delta R.T.: 0.000 min

Response: 632414

Conc: 18.21 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.211 min

Delta R.T.: 0.000 min

Response: 527283

Conc: 18.80 ng/ml