

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB120919\
 Data File : FB023169.D
 Signal(s) : FID2B.CH
 Acq On : 9 Dec 2019 17:22
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Dec 10 00:50:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
 Quant Title :
 QLast Update : Fri Nov 22 12:34:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	470156	19.140 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	864985	26.470 ng/ml
2) t 2,2,4-Trimethylpentane	7.416	1000218	28.684 ng/ml
3) t n-Heptane	7.756	324593	9.440 ng/ml
4) t Benzene	7.890	392369	9.417 ng/ml
6) t Toluene	10.629	1165698	28.937 ng/ml
7) t Ethylbenzene	13.069	350861	9.092 ng/ml
8) t m-Xylene	13.203	665126	18.278 ng/ml
9) t O-Xylene	13.930	641310	18.470 ng/ml
10) t 1,2,4-Trimethylbenzene	16.209	548062	19.543 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB120919\
Data File : FB023169.D
Signal(s) : FID2B.CH
Acq On : 9 Dec 2019 17:22
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Dec 10 00:50:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.689 min
Delta R.T.: -0.002 min
Response: 864985
Conc: 26.47 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min
Delta R.T.: -0.002 min
Response: 1000218
Conc: 28.68 ng/ml




#3 n-Heptane

R.T.: 7.756 min
Delta R.T.: -0.002 min
Response: 324593
Conc: 9.44 ng/ml



#4 Benzene


R.T.: 7.890 min
Delta R.T.: -0.002 min
Response: 392369
Conc: 9.42 ng/ml



#5 AAA-TFT


R.T.: 8.797 min
Delta R.T.: -0.002 min
Response: 470156
Conc: 19.14 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.629 min
Delta R.T.: -0.002 min
Response: 1165698
Conc: 28.94 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min
Delta R.T.: -0.001 min
Response: 350861
Conc: 9.09 ng/ml



#8 m-Xylene


R.T.: 13.203 min
Delta R.T.: 0.000 min
Response: 665126
Conc: 18.28 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: -0.001 min
Response: 641310
Conc: 18.47 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.209 min
Delta R.T.: -0.002 min
Response: 548062
Conc: 19.54 ng/ml