

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB120920\
 Data File : FB025316.D
 Signal(s) : FID2B.CH
 Acq On : 9 Dec 2020 17:48
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Dec 10 00:12:24 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120220.M
 Quant Title :
 QLast Update : Wed Dec 02 15:24:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.800	391419	18.619 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	843427	28.055 ng/ml
2) t 2,2,4-Trimethylpentane	7.418	848990	28.216 ng/ml
3) t n-Heptane	7.759	280381	9.468 ng/ml
4) t Benzene	7.894	350199	9.468 ng/ml
6) t Toluene	10.631	980168	27.893 ng/ml
7) t Ethylbenzene	13.071	298111	9.127 ng/ml
8) t m-Xylene	13.204	597391	18.457 ng/ml
9) t O-Xylene	13.932	574170	18.163 ng/ml
10) t 1,2,4-Trimethylbenzene	16.210	495030	18.230 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB120920\
Data File : FB025316.D
Signal(s) : FID2B.CH
Acq On : 9 Dec 2020 17:48
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Dec 10 00:12:24 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120220.M
Quant Title :
QLast Update : Wed Dec 02 15:24:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.688 min
Delta R.T.: 0.001 min
Response: 843427
Conc: 28.05 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 848990
Conc: 28.22 ng/ml




#3 n-Heptane

R.T.: 7.759 min
Delta R.T.: 0.001 min
Response: 280381
Conc: 9.47 ng/ml



#4 Benzene


R.T.: 7.894 min
Delta R.T.: 0.001 min
Response: 350199
Conc: 9.47 ng/ml



#5 AAA-TFT


R.T.: 8.800 min
Delta R.T.: 0.001 min
Response: 391419
Conc: 18.62 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.631 min
Delta R.T.: 0.001 min
Response: 980168
Conc: 27.89 ng/ml




#7 Ethylbenzene

R.T.: 13.071 min
Delta R.T.: 0.000 min
Response: 298111
Conc: 9.13 ng/ml



#8 m-Xylene


R.T.: 13.204 min
Delta R.T.: 0.000 min
Response: 597391
Conc: 18.46 ng/ml



#9 O-Xylene

R.T.: 13.932 min
Delta R.T.: 0.000 min
Response: 574170
Conc: 18.16 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.210 min
Delta R.T.: 0.000 min
Response: 495030
Conc: 18.23 ng/ml