

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\
 Method File : FB120718.M
 Title :
 Last Update : Fri Dec 07 17:57:38 2018
 Response Via : Initial Calibration

Calibration Files

= S1 =FB018969.D S2 =FB018965.D
 S3 =FB018966.D S4 =FB018967.D S5 =FB018968.D

Compound		S1	S2	S3	S4	S5	Avg	%RSD	
1) t	2-Methylpentane	1.408	1.503	1.476	1.702	1.555	1.529	E4	7.23
2) t	2,2,4-Trimeth...	1.423	1.527	1.484	1.707	1.562	1.541	E4	6.92
3) t	n-Heptane	1.282	1.560	1.449	1.651	1.490	1.486	E4	9.26
4) t	Benzene	1.484	1.823	1.786	2.047	1.865	1.801	E4	11.31
5) s	AAA-TFT	8.117	9.190	8.682	9.257	8.456	8.741	E3	5.55
6) t	Toluene	1.422	1.498	1.453	1.665	1.551	1.518	E4	6.31
7) t	Ethylbenzene	1.538	1.594	1.535	1.580	1.521	1.554	E4	2.02
8) t	m-Xylene	1.629	1.665	1.594	1.661	1.597	1.629	E4	2.07
9) t	o-Xylene	1.646	1.733	1.665	1.733	1.676	1.691	E4	2.38
10) t	1,2,4-Trimeth...	1.981	2.084	1.995	2.057	2.002	2.024	E4	2.19

(#) = Out of Range