

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB121118\
 Data File : FB019022.D
 Signal(s) : FID2B.CH
 Acq On : 11 Dec 2018 12:01
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Dec 12 04:32:03 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120718.M
 Quant Title :
 QLast Update : Fri Dec 07 17:57:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.111	142077	16.255 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.964	411804	26.937 ng/ml
2) t 2,2,4-Trimethylpentane	7.728	431434	28.004 ng/ml
3) t n-Heptane	8.067	135520	9.118 ng/ml
4) t Benzene	8.206	174145	9.668 ng/ml
6) t Toluene	10.960	445562	29.354 ng/ml
7) t Ethylbenzene	13.410	145478	9.364 ng/ml
8) t m-Xylene	13.543	309835	19.018 ng/ml
9) t O-Xylene	14.274	328059	19.402 ng/ml
10) t 1,2,4-Trimethylbenzene	16.557	386329	19.090 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB121118\
Data File : FB019022.D
Signal(s) : FID2B.CH
Acq On : 11 Dec 2018 12:01
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Dec 12 04:32:03 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120718.M
Quant Title :
QLast Update : Fri Dec 07 17:57:38 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.964 min
Delta R.T.: 0.003 min
Response: 411804
Conc: 26.94 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.728 min
Delta R.T.: 0.006 min
Response: 431434
Conc: 28.00 ng/ml




#3 n-Heptane

R.T.: 8.067 min
Delta R.T.: 0.004 min
Response: 135520
Conc: 9.12 ng/ml




#4 Benzene

R.T.: 8.206 min
Delta R.T.: 0.005 min
Response: 174145
Conc: 9.67 ng/ml




#5 AAA-TFT

R.T.: 9.111 min
Delta R.T.: 0.005 min
Response: 142077
Conc: 16.25 ng/ml




#6 Toluene

R.T.: 10.960 min
Delta R.T.: 0.005 min
Response: 445562
Conc: 29.35 ng/ml




#7 Ethylbenzene

R.T.: 13.410 min
Delta R.T.: 0.004 min
Response: 145478
Conc: 9.36 ng/ml




#8 m-Xylene

R.T.: 13.543 min
Delta R.T.: 0.005 min
Response: 309835
Conc: 19.02 ng/ml



#9 O-Xylene

R.T.: 14.274 min
Delta R.T.: 0.004 min
Response: 328059
Conc: 19.40 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.557 min
Delta R.T.: 0.004 min
Response: 386329
Conc: 19.09 ng/ml