

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB121520\
 Data File : FB025358.D
 Signal(s) : FID2B.CH
 Acq On : 15 Dec 2020 12:01
 Operator : DD/AJ
 Sample : L5069-04MS
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 KENS-B7-121120-13-15MS

Integration File: Calibration.e
 Quant Time: Dec 16 00:26:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120220.M
 Quant Title :
 QLast Update : Wed Dec 02 15:24:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.796	485010	23.070 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.682	923498	30.718 ng/ml
2) t 2,2,4-Trimethylpentane	7.413	951096	31.609 ng/ml
3) t n-Heptane	7.754	313345	10.582 ng/ml
4) t Benzene	7.889	383541	10.369 ng/ml
6) t Toluene	10.628	1088917	30.987 ng/ml
7) t Ethylbenzene	13.069	336783	10.311 ng/ml
8) t m-Xylene	13.202	672780	20.786 ng/ml
9) t O-Xylene	13.930	649617	20.549 ng/ml
10) t 1,2,4-Trimethylbenzene	16.207	563243	20.742 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB121520\
Data File : FB025358.D
Signal(s) : FID2B.CH
Acq On : 15 Dec 2020 12:01
Operator : DD/AJ
Sample : L5069-04MS
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
KENS-B7-121120-13-15MS

Integration File: Calibration.e
Quant Time: Dec 16 00:26:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120220.M
Quant Title :
QLast Update : Wed Dec 02 15:24:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.682 min
Delta R.T.: -0.005 min
Response: 923498
Conc: 30.72 ng/ml

Instrument :
FID_B
ClientSampleId :
KENS-B7-121120-13-15MS




#2 2,2,4-Trimethylpentane

R.T.: 7.413 min
Delta R.T.: -0.004 min
Response: 951096
Conc: 31.61 ng/ml




#3 n-Heptane

R.T.: 7.754 min
Delta R.T.: -0.004 min
Response: 313345
Conc: 10.58 ng/ml



#4 Benzene


R.T.: 7.889 min
Delta R.T.: -0.003 min
Response: 383541
Conc: 10.37 ng/ml



#5 AAA-TFT


R.T.: 8.796 min
Delta R.T.: -0.003 min
Response: 485010
Conc: 23.07 ng/ml

Instrument :
FID_B
ClientSampleId :
KENS-B7-121120-13-15MS




#6 Toluene

R.T.: 10.628 min
Delta R.T.: -0.002 min
Response: 1088917
Conc: 30.99 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min
Delta R.T.: -0.002 min
Response: 336783
Conc: 10.31 ng/ml



#8 m-Xylene


R.T.: 13.202 min
Delta R.T.: -0.002 min
Response: 672780
Conc: 20.79 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: -0.002 min
Response: 649617
Conc: 20.55 ng/ml

Instrument :
FID_B
ClientSampleId :
KENS-B7-121120-13-15MS



#10 1,2,4-Trimethylbenzene

R.T.: 16.207 min
Delta R.T.: -0.002 min
Response: 563243
Conc: 20.74 ng/ml