

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB121619\
 Data File : FB023240.D
 Signal(s) : FID2B.CH
 Acq On : 16 Dec 2019 19:41
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Dec 17 00:08:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
 Quant Title :
 QLast Update : Fri Nov 22 12:34:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.795	411185	16.739 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	755825	23.129 ng/ml
2) t 2,2,4-Trimethylpentane	7.413	875946	25.120 ng/ml
3) t n-Heptane	7.753	281804	8.196 ng/ml
4) t Benzene	7.889	364052	8.738 ng/ml
6) t Toluene	10.626	1040385	25.826 ng/ml
7) t Ethylbenzene	13.066	332657	8.621 ng/ml
8) t m-Xylene	13.199	632240	17.374 ng/ml
9) t O-Xylene	13.927	623504	17.958 ng/ml
10) t 1,2,4-Trimethylbenzene	16.206	533816	19.035 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB121619\
Data File : FB023240.D
Signal(s) : FID2B.CH
Acq On : 16 Dec 2019 19:41
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 18 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Dec 17 00:08:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.688 min
Delta R.T.: -0.003 min
Response: 755825
Conc: 23.13 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.413 min
Delta R.T.: -0.005 min
Response: 875946
Conc: 25.12 ng/ml




#3 n-Heptane

R.T.: 7.753 min
Delta R.T.: -0.004 min
Response: 281804
Conc: 8.20 ng/ml




#4 Benzene

R.T.: 7.889 min
Delta R.T.: -0.004 min
Response: 364052
Conc: 8.74 ng/ml




#5 AAA-TFT

R.T.: 8.795 min
Delta R.T.: -0.004 min
Response: 411185
Conc: 16.74 ng/ml




#6 Toluene

R.T.: 10.626 min
Delta R.T.: -0.004 min
Response: 1040385
Conc: 25.83 ng/ml




#7 Ethylbenzene

R.T.: 13.066 min
Delta R.T.: -0.004 min
Response: 332657
Conc: 8.62 ng/ml




#8 m-Xylene

R.T.: 13.199 min
Delta R.T.: -0.004 min
Response: 632240
Conc: 17.37 ng/ml



#9 O-Xylene

R.T.: 13.927 min
Delta R.T.: -0.004 min
Response: 623504
Conc: 17.96 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.206 min
Delta R.T.: -0.005 min
Response: 533816
Conc: 19.04 ng/ml