

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID_B\Data\FB121718\
Data File : FB019106.D
Signal(s) : FID2B.CH
Acq On : 17 Dec 2018 9:04
Operator : MJ/SJ
Sample : VBF1217S1
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Dec 17 14:38:54 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120718.M
Quant Title :
QLast Update : Fri Dec 07 17:57:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.112	166770	19.080 ng/ml
Target Compounds			
4) t Benzene	8.199	6779	0.376 ng/ml
6) t Toluene	10.977	4953	0.326 ng/ml
8) t m-Xylene	13.563	9240	0.567 ng/ml
10) t 1,2,4-Trimethylbenzene	16.536	345	0.017 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID_B\Data\FB121718\
Data File : FB019106.D
Signal(s) : FID2B.CH
Acq On : 17 Dec 2018 9:04
Operator : MJ/SJ
Sample : VBF1217S1
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Dec 17 14:38:54 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120718.M
Quant Title :
QLast Update : Fri Dec 07 17:57:38 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#4 Benzene

R.T.: 8.199 min
Delta R.T.: -0.002 min
Response: 6779
Conc: 0.38 ng/ml




#5 AAA-TFT

R.T.: 9.112 min
Delta R.T.: 0.005 min
Response: 166770
Conc: 19.08 ng/ml




#6 Toluene

R.T.: 10.977 min
Delta R.T.: 0.022 min
Response: 4953
Conc: 0.33 ng/ml



#8 m-Xylene

R.T.: 13.563 min
Delta R.T.: 0.025 min
Response: 9240
Conc: 0.57 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.536 min
Delta R.T.: -0.017 min
Response: 345
Conc: 0.02 ng/ml