

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB121719\
Data File : FB023259.D
Signal(s) : FID2B.CH
Acq On : 17 Dec 2019 13:51
Operator : DD/AJ
Sample : K6300-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
SSI-MW-2

Integration File: Calibration.e
Quant Time: Dec 18 00:58:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.798	540173	21.991 ng/ml
Target Compounds			
9) t O-Xylene	13.899	1489	0.043 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB121719\
Data File : FB023259.D
Signal(s) : FID2B.CH
Acq On : 17 Dec 2019 13:51
Operator : DD/AJ
Sample : K6300-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
SSI-MW-2

Integration File: Calibration.e
Quant Time: Dec 18 00:58:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT

R.T.: 8.798 min
Delta R.T.: -0.002 min
Response: 540173
Conc: 21.99 ng/ml

Instrument :
FID_B
ClientSampleId :
SSI-MW-2



#9 O-Xylene

R.T.: 13.899 min
Delta R.T.: -0.033 min
Response: 1489
Conc: 0.04 ng/ml