

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB121719\
 Data File : FB023251.D
 Signal(s) : FID2B.CH
 Acq On : 17 Dec 2019 9:30
 Operator : DD/AJ
 Sample : BSF1217S1
 Misc : 5.00g/5.00ml DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF1217S1

Integration File: Calibration.e
 Quant Time: Dec 18 00:56:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
 Quant Title :
 QLast Update : Fri Nov 22 12:34:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	497767	20.264 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	908570	27.804 ng/ml
2) t 2,2,4-Trimethylpentane	7.416	1025859	29.419 ng/ml
3) t n-Heptane	7.756	332633	9.674 ng/ml
4) t Benzene	7.890	399480	9.588 ng/ml
6) t Toluene	10.628	1157096	28.724 ng/ml
7) t Ethylbenzene	13.068	360449	9.341 ng/ml
8) t m-Xylene	13.201	689731	18.954 ng/ml
9) t O-Xylene	13.929	666970	19.210 ng/ml
10) t 1,2,4-Trimethylbenzene	16.207	558973	19.932 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB121719\
Data File : FB023251.D
Signal(s) : FID2B.CH
Acq On : 17 Dec 2019 9:30
Operator : DD/AJ
Sample : BSF1217S1
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF1217S1

Integration File: Calibration.e
Quant Time: Dec 18 00:56:52 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB112219.M
Quant Title :
QLast Update : Fri Nov 22 12:34:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.689 min
Delta R.T.: -0.002 min
Response: 908570
Conc: 27.80 ng/ml

Instrument :
FID_B
ClientSampleId :
BSF1217S1




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min
Delta R.T.: -0.002 min
Response: 1025859
Conc: 29.42 ng/ml




#3 n-Heptane

R.T.: 7.756 min
Delta R.T.: -0.002 min
Response: 332633
Conc: 9.67 ng/ml



#4 Benzene


R.T.: 7.890 min
Delta R.T.: -0.002 min
Response: 399480
Conc: 9.59 ng/ml



#5 AAA-TFT


R.T.: 8.797 min
Delta R.T.: -0.002 min
Response: 497767
Conc: 20.26 ng/ml

Instrument :
FID_B
ClientSampleId :
BSF1217S1




#6 Toluene

R.T.: 10.628 min
Delta R.T.: -0.002 min
Response: 1157096
Conc: 28.72 ng/ml




#7 Ethylbenzene

R.T.: 13.068 min
Delta R.T.: -0.003 min
Response: 360449
Conc: 9.34 ng/ml



#8 m-Xylene


R.T.: 13.201 min
Delta R.T.: -0.002 min
Response: 689731
Conc: 18.95 ng/ml



#9 O-Xylene

R.T.: 13.929 min
Delta R.T.: -0.002 min
Response: 666970
Conc: 19.21 ng/ml

Instrument :
FID_B
ClientSampleId :
BSF1217S1



#10 1,2,4-Trimethylbenzene

R.T.: 16.207 min
Delta R.T.: -0.003 min
Response: 558973
Conc: 19.93 ng/ml