

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB122324\
 Data File : FB031292.D
 Signal(s) : FID2B.CH
 Acq On : 23 Dec 2024 11:37
 Operator : YP/AJ
 Sample : VBF1223S1
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Dec 24 03:21:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121624.M
 Quant Title :
 QLast Update : Mon Dec 16 11:52:34 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.776	486655	18.446 ng/ml
Target Compounds			
6) t Toluene	10.612	7561	0.181 ng/ml
8) t m-Xylene	13.173	7419	0.186 ng/ml
9) t O-Xylene	13.904	7883	0.215 ng/ml
10) t 1,2,4-Trimethylbenzene	16.183	12044	0.404 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB122324\
Data File : FB031292.D
Signal(s) : FID2B.CH
Acq On : 23 Dec 2024 11:37
Operator : YP/AJ
Sample : VBF1223S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Dec 24 03:21:34 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121624.M
Quant Title :
QLast Update : Mon Dec 16 11:52:34 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT


R.T.: 8.776 min
Delta R.T.: -0.013 min
Response: 486655
Conc: 18.45 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.612 min
Delta R.T.: -0.004 min
Response: 7561
Conc: 0.18 ng/ml




#8 m-Xylene

R.T.: 13.173 min
Delta R.T.: -0.013 min
Response: 7419
Conc: 0.19 ng/ml



#9 O-Xylene

R.T.: 13.904 min
Delta R.T.: -0.009 min
Response: 7883
Conc: 0.21 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.183 min
Delta R.T.: -0.007 min
Response: 12044
Conc: 0.40 ng/ml

Instrument :
FID_B
ClientSampleId :